

Successful endoscopic ultrasound-guided hepaticogastrostomy with use of a novel drill dilator for challenging tract dilation


OPEN
ACCESS



► **Fig. 1** Photograph showing the novel drill dilator (Tornus ES; Asahi Intecc Co., Ltd.), which is advanced by clockwise rotation and withdrawn by counterclockwise rotation.



► **Fig. 2** Endoscopic ultrasound image showing a hypoechoic area around the intrahepatic bile duct because of the patient's frequent episodes of cholangitis (arrow), with the diameter of the intrahepatic bile duct being only 0.7 mm.



► **Video 1** Endoscopic ultrasound-guided hepaticogastrostomy is successfully performed following tract dilation using a novel drill dilator.

When endoscopic retrograde cholangiopancreatography (ERCP) fails, patients with benign biliary disease, such as common bile duct (CBD) stones or hepaticojunostomy stricture, are currently treated using the transhepatic approach after fistula creation by endoscopic ultrasound-guided hepaticogastrostomy (EUS-HGS) [1]. During EUS-HGS, tract dilation is required prior to stent deployment. Various devices have been reported for this purpose [2–4]; however, tract dilation can be challenging if the intrahepatic bile duct (IHBD) is not dilated and the bile duct wall is hard because of frequent epi-

sodes of cholangitis. A novel drill dilator (Tornus ES; Asahi Intecc Co., Ltd., Seto, Japan) that can overcome these difficulties has become available in Japan (► **Fig. 1**) [5]. This device guarantees tract dilation owing to its design based on screw mechanics. We describe a case of successful EUS-HGS in which the novel dilator was used for a challenging tract dilation. A 59-year-old man presented with frequent cholangitis due to CBD stones. As he had previously undergone distal gastrectomy with Roux-en-Y anastomosis for gastric cancer, EUS-HGS was considered to be an option. EUS revealed a

hypoechoic area around the IHBD owing to his frequent episodes of cholangitis (► **Fig. 2**). The diameter of the IHBD was 0.7 mm, and the IHBD was gently punctured using a 19G needle. After the successful puncture, contrast medium was injected. A 0.025-inch guidewire was then deployed into the CBD (► **Fig. 3 a**). Tract dilation was attempted with a balloon catheter, but the device could not be inserted into the biliary tract. Therefore, a Tornus dilator was inserted into the echoendoscope and was successfully advanced across the stomach and bile duct wall using clockwise rotation (► **Fig. 3 b**). After the tract had been dilated, an 8.5-Fr stent delivery system was successfully inserted and a metal stent was then deployed (► **Fig. 3 c**; ► **Video 1**). In conclusion, the Tornus ES dilator appears to be useful as a dilation device, particularly when tract dilation is difficult, as was the case in the patient described.

Endoscopy_UCTN_Code_TTT_1AS_2AD

Competing interests

The authors declare that they have no conflict of interest.



► **Fig. 3** Fluoroscopic images showing: **a** a 0.025-inch guidewire deployed after successful bile duct puncture; **b** tract dilation being performed using the novel drill dilator; **c** endoscopic ultrasound-guided hepaticogastrostomy successfully performed.

The authors

Takeshi Ogura, Yuki Uba, Masahiro Yamamura, Jyunichi Kawai, Hiroki Nishikawa
2nd Department of Internal Medicine, Osaka Medical and Pharmaceutical University, Osaka, Japan

Corresponding author

Takeshi Ogura, MD, PhD
2nd Department of Internal Medicine, Osaka Medical and Pharmaceutical University, 2-7 Daigakuchou, Takatsukishi, Osaka 569-8686, Japan
oguratakeshi0411@yahoo.co.jp

References

- [1] van Wanrooij RLJ, Bronswijk M, Kunda R et al. Therapeutic endoscopic ultrasound: European Society of Gastrointestinal Endoscopy (ESGE) Technical Review. *Endoscopy* 2022; 54: 310–332
- [2] Ogura T, Nakai Y, Iwashita T et al. Novel fine gauge electrocautery dilator for endoscopic ultrasound-guided biliary drainage: experimental and clinical evaluation study (with video). *Endosc Int Open* 2019; 7: E1652–E1657

- [3] Amano M, Ogura T, Onda S et al. Prospective clinical study of endoscopic ultrasound-guided biliary drainage using novel balloon catheter (with video). *J Gastroenterol Hepatol* 2017; 32: 716–720
- [4] Honjo M, Itoi T, Tsuchiya T et al. Safety and efficacy of ultra-tapered mechanical dilator for EUS-guided hepaticogastrostomy and pancreatic duct drainage compared with electrocautery dilator (with video). *Endosc Ultrasound* 2018; 7: 376–382
- [5] Yamada M, Hara K, Haba S et al. Endoscopic ultrasound-guided hepaticogastrostomy using a novel drill dilator. *Endoscopy* 2022. doi:10.1055/a-1838-3682

Bibliography

Endoscopy 2023; 55: E149–E150
DOI 10.1055/a-1956-0763
ISSN 0013-726X
published online 28.10.2022
© 2022. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution-NonDerivative-NonCommercial License, permitting copying and reproduction so long as the original work is given appropriate credit. Contents may not be used for commercial purposes, or adapted, remixed, transformed or built upon. (<https://creativecommons.org/licenses/by-nc-nd/4.0/>)
Georg Thieme Verlag KG, Rüdigerstraße 14, 70469 Stuttgart, Germany



ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>



Endoscopy E-Videos is an open access online section, reporting on interesting cases and new techniques in gastroenterological endoscopy. All papers include a high quality video and all contributions are freely accessible online. Processing charges apply (currently EUR 375), discounts and waivers acc. to HINARI are available.

This section has its own submission website at
<https://mc.manuscriptcentral.com/e-videos>