

Novel therapeutic thin endoscope facilitates endoscopic submucosal dissection for cervical esophageal cancer involving the pharyngo-esophageal junction




► **Fig. 1** Comparison of the novel therapeutic thin endoscope and normal diameter therapeutic endoscope. Left: Novel therapeutic thin endoscope has a 7.9-mm outer diameter and 3.2-mm working channel diameter. Right: Normal diameter therapeutic endoscope with a 9.8-mm outer diameter.



► **Fig. 2** Comparison of the downward angle of 160° (novel therapeutic thin endoscope) and 90° (conventional endoscope).

The cervical esophagus, especially the pharyngo-esophageal junction, is narrow; thus, the movement of the endoscope is severely limited [1,2]. This region also has strong flexion to the posterior wall; thus, it is occasionally difficult to approach a lesion with a conventional endoscope with an insufficient downward angle. Recently, a novel therapeutic thin endoscope, EG-840TP (Fujifilm, Tokyo, Japan), has been developed. In this endoscope, despite an outer diameter of 7.9mm, the working channel diameter is 3.2mm with a downward angle of




► **Fig. 3** Case description. **a** The lesion was 50 mm in size, occupied half the circumferential area, and was located on the posterior wall. **b** During the entire procedure, maneuverability and approachability were maintained well. **c** Wound of the endoscopic submucosal dissection. **d** En bloc resection was achieved. The procedure time was 100 min and the specimen size was 62 × 37 mm. The pathological diagnosis was squamous cell carcinoma, T1a, with a free horizontal and vertical margin.

160° (► **Fig. 1**, ► **Fig. 2**). Herein, we report a case in which this novel endoscope facilitates endoscopic submucosal dissection (ESD) for cervical esophageal cancer involving the pharyngo-esophageal junction (► **Video 1**).


A 58-year-old man was referred to our hospital for the treatment of cervical esophageal cancer. The lesion was 50mm in size, occupied a half circumferential region, and was located on the posterior wall, involving the pharyngo-esophageal junction. We decided to perform ESD. Tracheal intubation and general anesthesia without a laryngoscope were performed. The EG-840TP and DualKnifeJ (Olympus Medical Systems, Tokyo, Japan) were used. ESD was performed in underwater conditions ac-

ording to the following procedure: 1) mucosal incision and trimming of the distal side to create an endpoint, 2) a full circumferential incision, 3) dissection of both lateral edges of the submucosa, and 4) submucosal dissection of the central area [3]. During the entire procedure, maneuverability and approachability were maintained well. Finally, en bloc resection was achieved without any adverse events (► **Fig. 3**).

We were able to perform successful cervical esophageal ESD due to this novel endoscope. The thin outer diameter allows for free movement even in a narrow space and the sufficient downward angle improves approachability (► **Fig. 4**). This novel endoscope may open up new possibilities for endoscopic treatment by



► **Fig. 4** Illustration of cervical esophageal endoscopic submucosal dissection in conventional endoscope and novel therapeutic thin endoscope. **a** A conventional endoscope. The pharyngoesophageal junction is narrow and tortuous, thus maneuverability and approachability are poor. **b** Novel therapeutic thin endoscope. Thin outer diameter allows for free movement even in the narrow space (red circle). **c** Sufficient downward angle also improves the approachability (blue circle).



► **Video 1** Endoscopic submucosal dissection for cervical esophageal cancer involving the pharyngoesophageal junction using the novel therapeutic thin endoscope.

- [2] Iizuka T, Kikuchi D, Hoteya S et al. Efficacy and safety of endoscopic submucosal dissection for superficial cancer of the cervical esophagus. *Endosc Int Open* 2017; 5: E736–E741. doi:10.1055/s-0043-112493
- [3] Masunaga T, Kato M, Yahagi N. Successful endoscopic submucosal dissection using the water pressure method for cervical esophageal cancer. *Dig Endosc* 2021; 33: e93–e94. doi:10.1111/den.13973

Bibliography

Endoscopy 2023; 55: E602–E603
 DOI 10.1055/a-2044-0512
 ISSN 0013-726X
 © 2023. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution-NonDerivative-NonCommercial License, permitting copying and reproduction so long as the original work is given appropriate credit. Contents may not be used for commercial purposes, or adapted, remixed, transformed or built upon. (<https://creativecommons.org/licenses/by-nc-nd/4.0/>)

Georg Thieme Verlag KG, Rüdigerstraße 14,
 70469 Stuttgart, Germany



enabling movements that were not possible before.

Endoscopy_UCTN_Code_TTT_1AO_2AC

Competing interests

The authors declare that they have no conflict of interest.

The authors

Tepei Masunaga **Motohiko Kato** **Naohisa Yahagi**

Division of Research and Development for Minimally Invasive Treatment, Cancer Center, Keio University School of Medicine, Tokyo, Japan

Corresponding author

Motohiko Kato, MD

Division of Research and Development for Minimally Invasive Treatment, Cancer Center, Keio University, School of Medicine, 35 Shinanomachi, Shinjuku-ku, Tokyo, 160-8582, Japan
 motohikokato@keio.jp

References

- [1] Ariyoshi R, Toyonaga T, Tanaka S et al. Clinical outcomes of endoscopic submucosal dissection for superficial esophageal neoplasms extending to the cervical esophagus. *Endoscopy* 2018; 50: 613–617. doi:10.1055/s-0043-123761