



Successful retrieval of a fractured biliary guidewire using a newly developed endoscopic tapered sheath

OPEN
ACCESS



► **Fig. 1** Contrast-enhanced computed tomography scan showing a Bismuth–Corlette type II perihilar cholangiocarcinoma.



► **Fig. 2** Fluoroscopic images during endoscopic retrograde cholangiopancreatography showing: **a** a severe hilar stricture; **b** the guidewire fragment located beyond the malignant biliary obstruction.



► **Fig. 3** Fluoroscopic image showing the intraductal cholangioscopy forceps being inserted through the novel tapered endoscopic sheath to remove the guidewire fragment.

An 81-year-old man with Bismuth–Corlette type II perihilar cholangiocarcinoma (► **Fig. 1**) underwent endoscopic retrograde cholangiopancreatography (ERCP). Cholangiography showed a hilar stricture that was so severe that the biopsy forceps could not be passed alongside the tumor (► **Fig. 2 a**). During this ERCP, a guidewire was fractured. An endoscopic nasobiliary drainage tube was inserted into the right anterior duct, without removal of the fragment being attempted (► **Fig. 2 b**).


A second ERCP was subsequently performed and, after insertion of a 0.025-inch angle-tip guidewire (Visiglide2; Olympus Medical Systems), an ERCP catheter (MTW Endoskopie, Wesel, Germany) was placed over the guidewire fragment. An attempt to remove it using a basket catheter through the MTW catheter was however unsuccessful. As an alternative, a newly developed endoscopic tapered sheath (EndoSheather; Piolax, Inc., Kanagawa, Japan) was inserted through the hilar stricture, without dilation being required. The inner catheter and VisiGlide2 were then removed, leaving the outer sheath in position. Intraductal chol-



► **Fig. 4** Photograph of the removed guidewire fragment.

angiography forceps (SpyBite Max; Boston Scientific, Natick, Massachusetts, USA) were inserted through the outer sheath (► **Fig. 3**), which enabled the guidewire fragment finally to be removed (► **Fig. 4**; ► **Video 1**).

Guidewire fracture is a rare complication of ERCP. Although there are several reports of the removal of a fractured guidewire fragment, a fragment that is situated beyond a malignant biliary obstruction is considered difficult to remove [1–4]. We have described the first case in which a guidewire fragment that was beyond a malignant stricture was successfully removed using this novel tapered sheath and cholangioscopy forceps. Even with severe bile duct stenosis, as seen in this case, it appears possible to safely and



Video 1 A new technique is used to remove a fractured guidewire located in a peripheral bile duct beyond a malignant biliary obstruction using a newly developed endoscopic tapered sheath.

easily remove a fragment from a peripheral bile duct using this method. This new technique may in future become one of the standard methods for removal of guidewire fragments.

Endoscopy_UCTN_Code_TTT_1AR_2AK

Competing interests

The authors declare that they have no conflict of interest.

The authors

Yuki Mori¹, Akira Kurita²

- 1 Department of Gastroenterology and Hepatology, Medical Research Institute Kitano Hospital, PIIF Tazuke-Kofukai, Kitano, Osaka, Japan
- 2 Department of Gastroenterology and Hepatology, Rakuwakai Otowa Hospital, Kyoto, Japan

Corresponding author

Akira Kurita, MD, PhD

Department of Gastroenterology and Hepatology, Rakuwakai Otowa Hospital, 2 Otowachinji-cho, Yamashina-ku, Kyoto, 607-8062, Japan
kuritaaki1976@gmail.com

References

- [1] Fry LC, Linder JD, Monkemuller KE. Cholangitis as a result of hydrophilic guidewire fracture. *Gastrointest Endosc* 2002; 56: 943–944
- [2] Burdick JS, Schmalz MJ, Geenen JE. Guidewire fracture during endoscopic sphincterotomy. *Endoscopy* 1993; 25: 251–252
- [3] Nishikawa Y, Uza N, Yamauchi Y et al. Successful endoscopic removal of fractured guidewire fragments from a peripheral bile duct using a biliary stent delivery system and biopsy forceps. *Endoscopy* 2018; 50: E279–E280
- [4] Tanisaka Y, Ryozaawa S, Mizuide M et al. Fractured guidewire fragment on the peripheral side of a malignant biliary obstruction: successful removal using a biliary balloon dilation catheter. *Endoscopy* 2021; 53: E183–E184

Bibliography

Endoscopy 2023; 55: E706–E707

DOI 10.1055/a-2073-5147

ISSN 0013-726X

© 2023. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution License, permitting unrestricted use, distribution, and reproduction so long as the original work is properly cited.

(<https://creativecommons.org/licenses/by/4.0/>)

Georg Thieme Verlag KG, Rüdigerstraße 14, 70469 Stuttgart, Germany



ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>



E-Videos is an open access online section of the journal *Endoscopy*, reporting on interesting cases

and new techniques in gastroenterological endoscopy. All papers include a high-quality video and are published with a Creative Commons CC-BY license. Endoscopy E-Videos qualify for HINARI discounts and waivers and eligibility is automatically checked during the submission process. We grant 100% waivers to articles whose corresponding authors are based in Group A countries and 50% waivers to those who are based in Group B countries as classified by Research4Life (see: <https://www.research4life.org/access/eligibility/>).

This section has its own submission website at

<https://mc.manuscriptcentral.com/e-videos>