Successful endoscopic submucosal dissection for a cecal laterally spreading tumor using a motorized spiral enteroscope in a case with difficult intubation



In colorectal endoscopic submucosal dissection (ESD), it is essential to intubate with the endoscope in a straight position and maintain stable maneuverability during the procedure. It is however often difficult to obtain these good conditions, especially in patients with colon elongation, poor fixation, or severe obesity. Recently, a motorized spiral enteroscope (MSE) was launched [1–3] (► Fig. 1). MSE enables deep intubation and appropriate fixation by rotation of the fins around the scope to squeeze the intestinal tract. Here, we report a very challenging cecal ESD in which the MSE was extremely useful (> Video 1).

A severely obese 71-year-old man (body mass index 38 kg/m²) was found to have a 50-mm laterally spreading tumor in the cecum. Because cecal intubation was very difficult and ESD was considered very challenging, he was referred to our hospital. At first, we attempted ESD with a single-balloon endoscope, but we were unable to reach the lesion sufficiently after trying for more than 90 minutes. Therefore, we decided to perform ESD with an MSE.

We achieved successful cecal intubation in 13 minutes, without any difficulty. Thereafter, we performed ESD using the water pressure method [4,5] (> Fig. 2). Although it was difficult to approach the distal side of the lesion, we were able to overcome this difficulty by suctioning the air and appropriately using forward rotation. The stable maneuverability of the scope was maintained and the patient did not complain of discomfort during the procedure. Consistent with this, intraoperative abdominal radiography showed that there was no migration of gas into the downstream colon, and the straight scope position was maintained (> Fig. 3). En bloc resection was achieved in 62 minutes, without any adverse events (> Fig. 4).



▶ Fig. 1 Photographs showing the motorized spiral enteroscope with an energy device for colorectal endoscopic submucosal dissection delivered through the working channel: **a** full view; **b** close up of the tip (it can be equipped with a hood, can deliver energy devices, and has waterjet function capability).

MSE can overcome difficulties in challenging ESD procedures such as this by improving maneuverability, as well as aiding deep intubation into the small intestine, for which it was originally developed.

Endoscopy_UCTN_Code_TTT_1AO_2AG

Competing interests

The authors declare that they have no conflict of interest.











Fig.4 White-light photograph of the specimen stained with indigo carmine after successful R0 resection.

The authors

Kurato Miyazaki^{1,2}, Motohiko Kato² [©] Kaoru Takabayashi³, Noriko Matsuura² [©] Naohisa Yahagi²

- Division of Gastroenterology and Hepatology, Department of Internal Medicine, Keio University School of Medicine, Tokyo, Japan
- 2 Division of Research and Development for Minimally Invasive Treatment, Cancer Center, Keio University School of Medicine, Tokyo, Japan
- 3 Center for Diagnostic and Therapeutic Endoscopy, Keio University School of Medicine, Tokyo, Japan

Corresponding author

Motohiko Kato, MD, PhD

Division of Research and Development for Minimally Invasive Treatment, Cancer Center, Keio University School of Medicine, 35 Shinanomachi, Shinjuku-ku, Tokyo, 160-8582, Japan motohikokato@keio.jp



► Fig. 2 White-light image during endoscopic submucosal dissection using the water pressure method, which was possible because of the stability and maneuverability of the scope.

References

- Schneider M, Hollerich J, Beyna T. Deviceassisted enteroscopy: A review of available techniques and upcoming new technologies. World J Gastroenterol 2019; 25: 3538– 3545
- [2] Beyna T, Arvanitakis M, Schneider M et al. Motorised spiral enteroscopy: first prospective clinical feasibility study. Gut 2021; 70: 261–267
- [3] Al-Toma A, Oude Hergelink DM, Tenthof van Noorden J et al. Prospective evaluation of the motorized spiral enteroscope for previous incomplete colonoscopy. Endosc Int Open 2022; 10: E1112–E1117
- [4] Kato M, Takatori Y, Sasaki M et al. Water pressure method for duodenal endoscopic submucosal dissection (with video). Gastrointest Endosc 2021; 93: 942–949
- [5] Miyazaki K, Kato M, Kanai T et al. A successful case of endoscopic submucosal dissection using the water pressure method for early gastric cancer with severe fibrosis. VideoGIE 2022; 7: 219–222

Bibliography

Endoscopy 2023; 55: E717–E718 DOI 10.1055/a-2078-1165 ISSN 0013-726X © 2023. The Author(s). This is an open access article published by Thieme under the

terms of the Creative Commons Attribution License, permitting unrestricted use, distribution, and reproduction so long as the original work is properly cited. (https://creativecommons.org/licenses/by/4.0/) Georg Thieme Verlag KG, Rüdigerstraße 14, 70469 Stuttgart, Germany

