A new T-type electrosurgical knife with waterjet function used in probe mode: a safe technical variant for colorectal endoscopic submucosal dissection



The water-jet hydrodissection technique is an effective method for colorectal endoscopic submucosal dissection (ESD) [1], even during the technique learning curve [2]. The use of the T-type Hybrid-Knife in "probe mode" can facilitate ESD [3], even in complex situations [4, 5]. The design of the new HybridKnife Flex (**> Fig. 1**) may improve the precision and safety of the technique.

A 68-year-old woman was referred for ESD of a rectal polyp after screening colonoscopy. A 45-mm lesion, with granular nodular mixed laterally spreading tumor (LST) morphology, was found to be located 10 cm above the dentate line (> Fig. 2). Therapeutic endoscopy was performed using a T-type HybridKnife Flex of 1.5 mm, the ERBEJET system, a VIO3 electrosurgical unit (ERBE, Germany), and a colonoscope (Fujifilm, Japan) with transparent hood (Olympus, Japan) (> Video 1). In each phase of the ESD, we performed dynamic adjustment of the electrosurgical unit settings (> Fig. 3). The submucosal dissection was performed in preciseSECT mode with continuous activation (no "bumps"), allowing faster movement of the HybridKnife for stepwise dissection with enhanced hemostasis. The Flex design caliber is 2.6 mm and facilitates dissection with the endoscope being pushed without touching the knife, in a similar manner to painting on canvas ("brush technique"). The coagulation of blood vessels was carried out in probe mode using soft coagulation, approaching without any mechanical pressure, with continuous activation; vessels were subsequently cut with preciseSECT. The procedure time was 45 minutes. The resected specimen size was 65×50 mm (> Fig. 4). Histopathologic examination identified a tubular adenoma with intramucosal adenocarcinoma and free lateral and vertical resection margins.



Fig.1 Photographs and illustration of the design features of the new T-type HybridKnife Flex in probe mode, with a 0.3-mm depth and 1.2-mm contact surface.



► Fig. 2 Endoscopic image showing a 45-mm lesion with granular nodular mixed laterally spreading tumor morphology that was located 10 cm above the dentate line.

The new HybridKnife Flex used in probe mode may be a promising alternative technique for ESD as it significantly facilitates the precision of the technique. The flexibility and finesse of the electrode in combination with the electrocautery settings of the VIO3 electrosurgical unit



▶ Video 1 The new HybridKnife Flex is used for the removal of a rectal lesion. The submucosal dissection phase using preciseSECT mode with the HybridKnife in probe mode allows dissection to be performed using the "brush technique."

and water-jet hydrodissection technique could in future simplify ESD.

Endoscopy_UCTN_Code_TTT_1AQ_2AD_3AD



▶ Fig. 3 The dynamic adjustment of the electrosurgical unit settings is shown during the phases of endoscopic submucosal dissection: a thermocautery marks are made around the edge of the lesion using soft coagulation mode with the HybridKnife in probe mode; b to access the submucosa, endoCUT mode is used with the HybridKnife open to 1.5 mm; c the submucosal dissection phase is performed in preciseSECT mode with HybridKnife in probe mode (the Flex design has a diameter of 2.6 mm, which provides stability in the working channel of the endoscope, allowing dissection to be carried out by pushing the endoscope without having to touch the knife, in a manner similar to painting on canvas ["brush technique"]); d coagulation of blood vessels is carried out in probe mode using soft coagulation mode, with preciseSECT subsequently used to cut them.



Fig.4 Macroscopic appearance of the resected specimen.

Acknowledgement

We gratefully acknowledge the Department of Anesthesiology and Resuscitation for their invaluable assistance in the development of this technique. We are indebted to the endoscopy auxiliary staff for collaborating in this case. We express our appreciation to Dr. Fabiola Machés for her assistance in editing this manuscript.

Conflict of Interest

F. Ramos-Zabala is a consultant for Erbe España Soluciones Médicas. F.G. Rojo, J.G. López-Mañas, F.G. Sánchez, S.R. Serrado, M.G. Mayor, and A.A. Pérez declare that they have no conflict of interest.

The authors

Felipe Ramos-Zabala^{1,2}, Francisco J. Gallego Rojo³, Julio Guilarte López-Mañas⁴, Francisco Gallardo Sánchez³, Sara Reina Serrado³, Marian García-Mayor^{1,2}, Alejandra Alzina-Pérez^{1,2}

- 1 Gastroenterology, HM Monteprincipe University Hospital, Madrid, Spain
- 2 Clinical Sciences, Faculty of Medicine, Universidad San Pablo CEU, Madrid, Spain
- 3 Gastroenterology, Hospital de Poniente de Almeria, El Ejido, Spain
- 4 Gastroenterology, Hospital Comarcal de Baza, Baza, Spain

Corresponding author

Felipe Ramos-Zabala, MD, PhD

Servicio de Gastroenterología, Hospital Universitario HM Montepríncipe, HM Hospitales, Av. de Montepríncipe 25, 28660 Boadilla del Monte, Madrid, Spain framoshdiaz@gmail.com

References

- Repici A, Hassan C, Pagano N et al. High efficacy of endoscopic submucosal dissection for rectal laterally spreading tumors larger than 3 cm. Gastrointest Endosc 2013; 77: 96–101
- [2] Ramos-Zabala F, Parra-Blanco A, Beg S et al. Feasibility and learning curve of unsupervised colorectal endoscopic submucosal hydrodissection at a Western Center. Eur J Gastroenterol Hepatol 2020; 32: 804–812. doi:10.1097/MEG.000000000001703
- [3] Ramos-Zabala F, Garcia-Mayor M, Dominguez-Pino A et al. Combination of immersion in saline solution, pocket-creation method, water-jet hydrodissection, and hybrid knife "probe mode" simplifies endoscopic submucosal dissection in giant rectal polyp. VideoGIE 2019; 4: 478–480
- [4] Ramos-Zabala F, Beg S, García-Mayor M et al. Novel approach to endoscopic submucosal dissection of a cecal lesion with non-lifting sign by submucosal fatty tissue using selective-regulation high-pressure water-jet method and immersion in saline solution. VideoGIE 2020; 5: 116–119
- [5] Ramos-Zabala F, Gil-Páez C, Alzina-Pérez A et al. "Trans-tattoo in immersion" method for the removal of a recurrent, previously tattooed adenoma using endoscopic submucosal hydrodissection. Endoscopy 2020; 52: E408–E410. doi:10.1055/a-1147-1206

Bibliography

Endoscopy 2024; 56: E532–E533 DOI 10.1055/a-2336-0941

ISSN 0013-726X

© 2024. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution License, permitting unrestricted use, distribution, and reproduction so long as the original work is properly cited.

(https://creativecommons.org/licenses/by/4.0/) Georg Thieme Verlag KG, Rüdigerstraße 14,

70469 Stuttgart, Germany

