



## A novel technique for peroral direct cholangioscopy




**Fig. 1** A typical bend in both the flexible and stiff end of a stiff wire.



**Fig. 2** The nasendoscope in position in the common bile duct (CBD) with the pre-bent wire still in place (arrow). The wire is subsequently removed to allow biopsy and flushing.



**Fig. 3** Second-order ducts, stone remnants, and mucus.



**Fig. 4** A ductal stone still in place following a balloon trawl.



**Fig. 5** Biopsy forceps in position.

Direct cholangioscopy has been possible since the 1970s [1] but the reliability of the intubation of the common bile duct (CBD) and second-order ducts has limited the applicability of the technique. One of the limitations to accessing the ducts is the acute angulation between the duodenum and the CBD [2]. Various methods

have been employed to access the ducts including mother-and-baby endoscopes and balloon assisted intubation [3]. Mother-and-baby systems are expensive, require additional equipment, and provide poorer quality images than standard endoscopes. Balloon-assisted intubation carries a risk of damage to the CBD, and requires special

equipment with an associated learning curve [4].

We have developed a novel technique that involves the use of a standard stiff wire (SMGW, Marflow AG, Switzerland) bent to the angulation of the duodenum-CBD junction (▶ **Fig. 1**) and passed through the working channel of a standard 5.6-mm Olympus nasendoscope. The technique begins after endoscopic retrograde cholangiopancreatography (ERCP) and sphincterotomy with a nasendoscope passed orally into the duodenum with the tip directed at the sphincter of Oddi. A pre-bent stiff wire is passed through the working channel until the bend in the wire sits within the flexible portion of the endoscope. The nasendoscope is then railroaded over the pre-bent wire into the CBD by advancing the nasendoscope whilst holding the wire fixed (▶ **Fig. 2**). Once the flexible portion of the nasendoscope is completely within the CBD, the wire can be removed to allow full use of the working channel.

Using this technique, we have successfully intubated second-order ducts (▶ **Fig. 3**) and have been able to make reliable in vivo diagnosis using high definition endoscopes, remove difficult stones (▶ **Fig. 4**), and biopsy lesions (▶ **Fig. 5**) under direct vision without complications. The advantages of our technique are its low cost, that no additional equipment is required, and that it has a short learning curve. More research is needed into the technique to determine its limits and the associated risks.

Endoscopy\_UCTN\_Code\_TTT\_1AR\_2AB

**Competing interests:** None

**Jim L. Callaghan, Gaius Longcroft-Wheaton, Andrew J. Fowell, Richard D. Ellis, Pradeep Bhandari, Patrick M. Goggin**

Department of Gastroenterology, Queen Alexandra's Hospital, Cosham, Portsmouth, PO6 3LY, UK

### References

- 1 *Urakami Y.* Peroral cholangiopancreatography (PCPS) and peroral direct cholangioscopy (PDCS). *Endoscopy* 1980; 12: 30–37
- 2 *Waxman I, Dillon T, Chmura K et al.* Feasibility of a novel system for intraductal balloon-anchored direct peroral cholangioscopy and endotherapy with an ultraslim endoscope (with videos). *Gastrointest Endosc* 2010; 72: 1052–1056

- 3 Moon JH, Ko BM, Choi HJ et al. Intraductal balloon-guided direct peroral cholangioscopy with an ultraslim upper endoscope (with videos). *Gastrointest Endosc* 2009; 70: 297–302
- 4 Pohl J, Ell C. Direct transnasal cholangioscopy with ultraslim endoscopes: a one-step intraductal balloon-guided approach. *Gastrointest Endosc* 2011; 74: 309–316

**Bibliography**

**DOI** <http://dx.doi.org/10.1055/s-0034-1390921>  
Endoscopy 2014; 46: E672–E673  
© Georg Thieme Verlag KG  
Stuttgart · New York  
ISSN 0013-726X

**Corresponding author**

**Patrick M. Gogglin, MD**  
Queen Alexandra's Hospital –  
Luminal Gastroenterology  
Southwick Hill Road  
Cosham  
Portsmouth PO6 3LY  
UK  
Fax: +44-23-92286303  
[Patrick.gogglin@porthosp.nhs.uk](mailto:Patrick.gogglin@porthosp.nhs.uk)