THIEME

©<u>()</u> = S

E-Procurement of Medical Laboratory Equipment: Experiences from a Tertiary Care Oncology Center in India

Naveen Bansal¹ Charu Singh² Sankalp Sancheti³

¹ Department of Transfusion Medicine, Homi Bhabha Cancer Hospital and Research Centre, Chandigarh, Punjab, India

² Department of Microbiology, Homi Bhabha Cancer Hospital and Research Centre, Chandigarh, Punjab, India

³ Department of Oncopathology, Homi Bhabha Cancer Hospital and Research Centre, Chandigarh, Punjab, India

⁴ Department of Surgical Oncology, Homi Bhabha Cancer Hospital and Research Centre, Chandigarh, Punjab, India

South Asian J Cancer

Dear Sir,

Patient outcomes and healthcare delivery are directly linked to the availability of calibrated, well-functioning, and quality laboratory medical equipment.¹ Quality assurance in medical laboratory is also directly dependent upon the equipment.² In low- and middle-income countries (LMIC), indiscriminate procurement methods among others are responsible for nonfunctioning of 40 to 70% of the medical equipment.³ Public procurement amounts to almost 20% of the gross domestic production of India.⁴ The Government e Marketplace (GeM) is an online public procurement platform launched in the year 2016 by Government of India with an aim to create a transparent procurement portal for government organizations.⁵ GeM has various advantages including (i) ease of purchase, (ii) transparency, and (iii) time bound delivery of goods. However, procurement of medical devices and equipment is a critical issue and has its own challenges.⁶ Our hospital is an upcoming government tertiary care oncology center with dedicated departments of oncopathology, transfusion medicine, transplant immunology, and microbiology. All the departments are being equipped with advanced state-of-the art medical equipment. We wish to share the issues identified and their possible solutions based on the experiences of e-procurement of medical laboratory equipment through GeM for our hospital.

Naveen Bansal

Issue 1: Quality Assurance

Ashish Gulia⁴⁰

Address for correspondence

140901, Punjab, India

Sankalp Sancheti, MD, Professor,

Homi Bhabha Cancer Hospital and

Research Centre, New Chandigarh

(e-mail: sankalpsancheti123@rediff.com).

Department of Oncopathology,

The equipment upon delivery to the hospital were jointly inspected by the biomedical engineering department and the end-user department. It was observed that many a times, the technical specifications of the equipment upon physical verification were not matching the technical specifications of the equipment as declared on GeM portal. Moreover, it was also observed that, in many cases, the equipment were not Conformite Europeenne/United States Food and Drug Administration/ International Organization for Standardization certified, whereas on GeM the vendor had declared these certifications. Thus, these equipment had to be rejected, which led to delay in the procurement process. The results are similar to study findings by Mehra et al in which a discrepancy of specifications resulted in delay or cancellation of the procurement process.⁷

Solution: A specialized department dealing with medical and surgical items should be created in the GeM handling government body. The department should have personnel with biomedical engineering and medical background. It has been demonstrated that biomedical engineering can play a very important role in quality assurance of healthcare technology in LMIC.⁸ The specialized department shall scrutinize the documents and certifications of vendors, before enlisting

DOI https://doi.org/10.1055/s-0043-1771443 ISSN 2278-330X

How to cite this article: Bansal N, Singh C, Sancheti S, et al. E-Procurement of Medical Laboratory Equipment: Experiences from a Tertiary Care Oncology Center in India. South Asian J Cancer 2023;00(00):00–00 © 2023. MedIntel Services Pvt Ltd. All rights reserved.

This is an open access article published by Thieme under the terms of the Creative Commons Attribution-NonDerivative-NonCommercial-License, permitting copying and reproduction so long as the original work is given appropriate credit. Contents may not be used for commercial purposes, or adapted, remixed, transformed or built upon. (https://creativecommons.org/licenses/by-nc-nd/ 4.0/)

Thieme Medical and Scientific Publishers Pvt. Ltd., A-12, 2nd Floor, Sector 2, Noida-201301 UP, India

them on GeM portal. The sellers/vendors should understand that the quality of medical equipment directly affects the patient healthcare delivery and therefore they have an ethical responsibility to deliver quality healthcare equipment as declared on the GeM.

Issue 2: Servicing/Repair of the Equipment

The downtime of the medical equipment is a critical factor in effective healthcare delivery and must be kept at minimum.⁹ It was observed that the medical equipment on GeM only come with a time bound warranty. Sometimes, the seller is located far from user location. Hence, effectively the onus is on the end-user to minimize the downtime in case of any equipment breakdown.

Solution: A clause to minimize downtime in case of equipment breakdown should be available for medical equipment.

To conclude, GeM is an effective and transparent tool for public procurement. However, as far as procurement process of medical equipment is concerned, various improvements need to be done in the GeM portal for streamlining of the procurement process.

Conflicts of Interest None declared.

References

- 1 Diaconu K, Chen YF, Cummins C, Jimenez Moyao G, Manaseki-Holland S, Lilford R. Methods for medical device and equipment procurement and prioritization within low- and middle-income countries: findings of a systematic literature review. Global Health 2017;13(01):59 Published 2017 Aug 18. Doi: 10.1186/ s12992-017-0280-2
- 2 Garraud O, Schneider T. International collaboration for blood safety: the French-African experience. Transfus Clin Biol 2021; 28(02):154–157
- 3 World Health Organization. Medical Device Regulations: Global Overview and Guiding Principles. Geneva2003
- 4 Mehta PS. Government procurement in India: domestic regulations and trade prospects. Jaipur, India: CUTS international; 2012. Accessed on June 23, 2023 at: http://www.cuts-citee.org/pdf/Government-Procurement-in-India_Domestic-Regulations-Trade-Prospects.pdf
- 5 Devendra P. Government E-Marketplace: a gem of an idea. Int J Emerging Technol Innov Res 2019;6(04):192–196
- 6 Ventola CL. Challenges in evaluating and standardizing medical devices in health care facilities. World Heath Organization, P&T 2008;33(06):348–359
- 7 Mehra A, Sharma S, Koushal V. Implementation of government emarketplace in a tertiary care hospital of North India: advantages and challenges. Int J Res Med Sci 2020;8:3934–3937
- 8 Cook DA. A protocol for the measurement of downtime of medical equipment. Br J Radiol 1997;70:279–290
- 9 De Maria C, Díaz Lantada A, Jämsä T, Pecchia L, Ahluwalia A. Biomedical engineering in low- and middle-income settings: analysis of current state, challenges and best practices. Health Technol (Berl) 2022;12(03):643–653