

Advances in artificial intelligence and computer science for computer-aided diagnosis of colorectal polyps: current status



Authors

Querijn NE van Bokhorst^{1,2}, Britt BSL Houwen^{1,2}, Yark Hazewinkel³, Paul Fockens^{1,2}, Evelien Dekker^{1,2}

Institutions

- 1 Department of Gastroenterology and Hepatology, Amsterdam University Medical Centers, location Academic Medical Center, Amsterdam, the Netherlands
- 2 Amsterdam Gastroenterology Endocrinology Metabolism, Amsterdam, the Netherlands
- 3 Department of Gastroenterology and Hepatology, Tergooi Medical Center, Hilversum, the Netherlands

received 4.2.2023

accepted after revision 8.5.2023

Bibliography

Endosc Int Open 2023; 11: E752–E767

DOI 10.1055/a-2098-1999

ISSN 2364-3722

© 2023. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution License, permitting unrestricted use, distribution, and reproduction so long as the original work is properly cited. (<https://creativecommons.org/licenses/by/4.0/>)

Georg Thieme Verlag KG, Rüdigerstraße 14,
70469 Stuttgart, Germany

Additional material is available at
<https://doi.org/10.1055/a-2098-1999>

Corresponding author

Evelien Dekker, MD PhD, Amsterdam University Medical Centers, location Academic Medical Center, Department of Gastroenterology and Hepatology, Meibergdreef 9, 1105 AZ Amsterdam, The Netherlands
e.dekker@amsterdamumc.nl

ABSTRACT

Colonoscopy is considered the gold standard for detection of colorectal cancer and its precursor lesions. However, colonoscopy outcomes may differ depending on the endoscopist performing the procedure. Among others, this relates to differences in ability of endoscopists to accurately assess polyp characteristics that are essential for clinical decision making. These characteristics concern polyp location, size and morphology, as well as several histological polyp features that can be predicted based on polyp phenotype. Polyp assessment with aid of computer-aided diagnosis (CADx) systems might provide opportunities to optimize general polyp assessment accuracy. However, a broad overview of available studies concerning performance of CADx systems for diagnosis of different polyp characteristics and histological features is lacking. Hence, within this narrative review we aimed to provide such an overview. We highlight that most significant advancements in the field of computer-aided polyp assessment involve systems for optical differentiation between neoplastic and non-neoplastic lesions, with several studies showing the ability of such systems to perform at expert levels in real-time clinical settings. With commercial availability of some of these systems, first steps towards improved endoscopy quality with use of CADx systems in daily practice might be ahead. However, development of CADx systems for assessment of polyp characteristics size and location, as well as prediction of degree of dysplasia and invasion depth, are still in more preliminary stages while evaluation of these systems in real-time clinical settings is still warranted. Moreover, computer-aided diagnosis of polyp morphology is a field yet to be explored.

Introduction

Colorectal cancer (CRC) is the third most commonly diagnosed malignancy and the second leading cause of cancer related-death in the world [1]. CRC develops from precancerous polyps through several (epi)genetic pathways [2]. Colonoscopy is con-

sidered the gold standard for detection and diagnosis of CRC and its precursor lesions [3, 4]. Moreover, colonoscopy provides opportunities for endoscopic resection of precancerous polyps, which is known to be effective to prevent CRC [5].

Despite reported benefits, colonoscopy outcomes may differ depending on quality of the endoscopist performing the procedure. Among others, this relates to differences in ability of endoscopists to accurately assess polyp characteristics such as location [6], size [7, 8, 9, 10] and morphology [11, 12, 13], as well as differences in their performance in predicting histological polyp features (e.g. histological subtype [14, 15, 16], grade of dysplasia [17, 18, 19], and, in case of suspected malignancy, presence of deep submucosal invasion [DSI] [20, 21, 22]). These polyp characteristics are essential to decide on the indication for resection and histopathological analysis [23, 24], appropriate resection method [25, 26] and appropriate surveillance interval [27, 28]. Hence, inaccurate endoscopic polyp assessment could lead to higher patient and economic burden due to unnecessary polyp resection and analysis, as well as suboptimal treatment and/or surveillance regimens.

Over the last decade, artificial intelligence (AI) in biomedical science has received growing attention. AI can be defined as the simulation of human intelligence by computer systems [29]. Specific AI techniques such as machine learning can be used to make machines (computers) smarter through experience-based learning [30, 31]. Since computer systems can be trained with a large amount of high quality and expert-annotated data, they could possibly serve as an objective, real-time, expert-level second observer modality during colonoscopy procedures. This might provide opportunities to reduce interobserver variability and improve general polyp assessment accuracy.

While evidence is currently scattered, we aimed to write a narrative review to provide a broad overview of current developments within the field of AI and computer science for computer-aided assessment of colorectal polyps. This includes assessment of polyp location, size, morphology and histology, including degree of dysplasia (low grade dysplasia [LGD] versus high grade dysplasia [HGD]) and, in case of suspected malignancy, invasion depth. Since computer-aided polyp detection concerns an already more thoroughly studied and evaluated topic [32, 33], developments within this field will not be addressed within this review.

Methods

A comprehensive literature search was performed in the MEDLINE/PubMed, Embase and Cochrane Libraries from the inception of the databases up to and including the 17th of July 2022. Key search terms used were “colorectal,” “polyp,” “artificial intelligence,” “size,” “location,” “morphology,” “histology,” “dysplasia” and “invasion depth.” Only studies published in English were screened. Reference lists of retrieved studies were manually screened to identify other relevant publications.

Results

Computer-aided assessment of polyp location

Accurate determination of polyp location is important to facilitate identification of a polyp or polypectomy site during consecutive colonoscopies and/or surgical procedures. In addition,

polyp location can aid in polyp histology prediction [34] and is important to adopt the ‘leave-in-situ’ optical diagnosis strategy in daily practice [23, 24].

To determine the location of the endoscope tip during colonoscopy procedures, and hence the location of observed polyps, endoscopists often rely on identification of various endoscopic anatomical landmarks and differences in colonic caliber, color tones and vasculature of different colon segments [35]. Endoscope intubation depth in centimeters could also be used. However, due to variations in colon length, shape and anatomy [36, 37, 38], change in colon length and position due to insufflation and endoscope intubation, and curving and bending of the endoscope due to the colon’s flexibility and elasticity, the accuracy of these methods seems limited. This is illustrated by earlier studies describing considerable interobserver variability [6] and 18% to 34% incorrect endoscopic localization of colorectal lesions when compared to findings during consecutive surgical procedures [39, 40, 41, 42, 43, 44].

Several deep learning approaches for orientation in the colon based on analysis of endoscopic videos and images have been proposed (► **Table 1**) [45, 46, 47]. Two studies described deep learning approaches for either recognition of anatomical landmarks [45] or distinguishing different colon segments [46] (accuracies 66.6% to 92.0%). Another study described several camera localization approaches, among which the localization approach based on analysis of camera motion in between colonoscopy video frames reached highest accuracy (71.8% in test set) [47].

Proposed systems could possibly aid endoscopists in orientation within the colon. However, current studies still concern feasibility studies and accuracy is mostly still limited. Besides, usage of a segment classification that assumes that all colons and segments are of similar length currently limits feasibility of the proposed motion-based localization system [47]. In addition, the issue concerning the lack of a solid reference standard should be addressed. While mostly only estimation of position within the colon by the endoscopist is available as reference standard, some sort of bias concerning training and (clinical) validation of such systems will likely always be present.

Toward the future, the issue of a lack of a solid reference standard could possibly be addressed by using magnetic endoscopic imaging (MEI) devices. These devices can improve accuracy of determination of location within the colon during colonoscopy [39, 48, 49, 50]. However, performance with aid of MEI devices is also not flawless and large-scale clinical trials assessing specific benefits of these devices for improving accuracy of polyp localization are still scarce. Thus, there is a need for further optimization and validation of MEI-assisted localization approaches, which may also improve the feasibility of existing deep learning approaches based on MEI data and images [47, 51]. Simultaneously, composition of more robust datasets for algorithm training, preferably only containing images/videos that are annotated by multiple experts, could aid in creating a more reliable reference standard. Variability in colon length could possibly be assessed, and accounted for, by using recently developed applications for image depth estimation and topographical reconstruction [52, 53, 54], assessment of endoscope

► Table 1 Overview of studies describing deep learning approaches for determination of location within the colon based on analysis of endoscopic videos and images.

| | Year | Described approach | Classification groups | Datasets* | Results Accuracy (%) |
|-------------------|------|--|---|--|----------------------|
| Che et al. [45] | 2021 | Deep learning model for recognition of endoscopic anatomical landmarks within video-derived colonoscopy images | Hepatic flexure Splenic flexure SDCJ | Training set: 6,911 images Test set: 1,729 images | 90.7–92.0† |
| Saito et al. [46] | 2021 | Deep learning model for distinguishing endoscopic colorectal images captured within different segments of the colon | Terminal ileum Cecum ACTTC DCTSC Rectum Anus | Training set: 9,995 images Test set: 5,121 images | 66.6 |
| Yao et al. [47] | 2021 | Deep learning model for estimation of relative location of the endoscope camera within the colon based on (analysis of) camera motion in between video frames‡ | Cecum Ascending colon Transverse colon Descending colon Sigmoid Rectum | Training set: 13 videos Test set: 3 videos | 71.8 |

SDCJ, sigmoid-descending colon junction; ACTTC, ascending colon to transverse colon; DCTSC, descending colon to sigmoid colon

*Data used for internal validation is reported as part of the training set.

†After post-processing through identification of incorrectly predicted frames (based on their temporal distribution) and reassigning these frames to the correct class, accuracies increased up to 99.8%.

‡Results for other methods (based on withdrawal time analysis, based on endoscope imaging device) not reported due to inferior results.

camera pose [55] and endoscopic three-dimensional (3D) colon reconstruction [56, 57, 58, 59]. Besides, 3D colon reconstruction [56, 57, 58, 59] techniques might open doors for development of other polyp localization approaches, since these could potentially visualize detected polyps within reconstructions of the complete colon.

Computer-aided assessment of polyp size

Polyp size has been shown to be associated with the risk that a polyp harbors advanced histological features [60], as well as the risk of metachronous advanced lesions and CRC [27, 28]. Hence, recommendations for appropriate resection method [25, 26] and surveillance intervals [27, 28] are determined, among other factors, by polyp size. Besides, polyp size determines whether a polyp can be included in the 'leave-in-situ' and 'resect-and-discard' optical diagnosis strategies for diminutive (1 to 5 mm) polyps [23, 24].

In daily practice, polyp size is based on visual estimation by the endoscopist. However, this strategy is prone to interobserver variability [7, 8, 9, 10], resulting in 10% to 35% inappropriate surveillance recommendations [9, 10]. To reduce interobserver variability, methods for automated polyp size measurement using deep learning approaches [61, 62, 63, 64, 65] and computer vision techniques [64, 66] have been proposed (► **Table 2**). Reported accuracies within these studies ranged between 79.2% to 88.0% [61, 62, 64]. Two studies benchmarking the performance of computer systems against that of endoscopists showed that computer systems may reach superior accuracy [64, 65].

While most studies showed promising results, some issues should be addressed. Most importantly, similar to polyp location, a robust reference standard is not available for polyp size.

This is illustrated by the fact that different reference standards were used in the different studies, limiting robustness and comparison of performance of the different systems. Additionally, several studies used binary polyp size classifications [61, 62, 64]. Use of binary approaches hampers reliable comparison to systems using exact polyp size estimation approaches.

Next steps for the development of more robust computerized polyp size measurement methods should include prospective evaluation of proposed systems in real-time clinical settings. Simultaneously, the problem concerning the lack of a solid reference standard might possibly be addressed through usage of recently developed endoscope-integrated or -attached polyp measurement tools [67, 68, 69]. Although it is unlikely that these tools will facilitate determination of the true size of polyps without a certain margin of error, as can only be accomplished by measuring polyps in colon (segment) resection specimens, they could possibly aid in obtaining highly reliable estimates of polyp size within in vivo settings. This relates to the fact that these tools can be precisely calibrated and validated using (artificial) polyps of known size in ex vivo settings. However, in order to gain further insights into feasibility of these tools, large-scale clinical studies validating accuracy of these tools are still required.

Computer-aided assessment of polyp morphology

Polyp morphology is an important feature for polyp malignancy risk-assessment [70] and can aid endoscopists in prediction of presence of DSI [20, 21, 71, 72]. As such, morphology also aids in selecting the optimal resection method [25, 26]. Assessment of polyp morphology is usually performed based on the Paris classification system [73] or laterally spreading tumor classification [74], but accuracy is known to be observer-dependent

► **Table 2** Overview of studies describing deep learning approaches or computer vision techniques for endoscopic polyp size measurement.

| | Year | Described approach or technique | Classification groups | Dataset(s)* | Size ground truth | Endoscopist comparison group (experience) | Measurement method comparison group | Results | |
|------------------------|------|---|--------------------------------------|---|--|--|---|-------------------|---------------------------|
| | | | | | | | | CADx (accuracy %) | Endoscopists (accuracy %) |
| Chadebecq et al. [66] | 2015 | Detection of Infocus-Break-point | Exact size estimation | Training set: 15 colonoscopy videos Test set: 5 colonoscopy videos | Visual estimation endoscopists (surgical tool as reference) | N/A | N/A | N/A [†] | N/A |
| Itoh et al. [61] | 2018 | Deep learning model | Binary approach: ≤ 10 mm vs. ≥ 10 mm | Training set: 34,396 images Test set: 13,093 images | Unspecified | N/A | N/A | 79.2 | N/A |
| Itoh et al. [62] | 2021 | Deep learning model | Binary approach: ≤ 10 mm vs. ≥ 10 mm | Training set: 94,980 images Test set: 15,569 images | Measurement with sheath of polypectomy snare as reference (consensus of 3 experts) | N/A | N/A | 81.0–88.0 | N/A |
| Su et al. [63] | 2021 | Deep learning model | Exact size estimation | Training set: N/A Test set: N/A | Pre-measured balls used for model development | N/A | N/A | N/A [‡] | N/A |
| Abdelrahim et al. [64] | 2022 | Photogrammetric imaging (structure from motion) technique | Binary approach: ≤ 5 mm vs. ≥ 5 mm | Training set: not reported Test set: 22 videos | Phantom polyps of known size | 10 endoscopists (varying degree of experience) | Visual estimation | 85.2 | 59.9 [§] |
| | | Deep learning model | Binary approach: ≤ 5 mm vs. ≥ 5 mm | Training set: 219 videos Test set: 10 videos | Visual size estimation endoscopists (mean of 3 experts) | N/A | N/A | 80.0 | N/A |
| Kwak et al. [65] | 2022 | Deep learning model | Exact size estimation | Training set: N/A [¶] Test set: 90 images | Measurement with ruler after resection | 4 experts (> 10,000 colonoscopies), 4 trainees (< 200 colonoscopies) | Visual estimation, opened snare measurement | N/A ^{††} | N/A ^{‡‡§§} |

CADx, computer-aided diagnosis; N/A, not available.

* Data used for internal validation reported as part of the training set.

† Instead of accuracy, mean error from ground truth reported: 4.5% to 6.4% (≈0.2 to 0.3 mm).

‡ Study described the process of model development for polyp size estimation. No specific results in terms of accuracy, sensitivity, specificity, negative predictive value, and positive predictive value reported.

§ Significant differences compared to CADx performance ($P < 0.05$).

¶ Model was built based on four datasets that are widely used for retinal vascular segmentation research. No specific polyp images were used for training.

†† Instead of accuracy concordance correlation coefficient (CCC) reported: 0.961.

‡‡ Instead of accuracy concordance correlation coefficient (CCC) reported: for visual size estimation CCC ranged between 0.650 and 0.758 (experts) and 0.465 and 0.703 (trainees). For open biopsy forceps size estimation CCC ranged between 0.789 and 0.815 (experts) and 0.657 and 0.762 (trainees).

§§ For visual size estimation significant difference reported for all endoscopists. For open biopsy forceps measurement significant differences reported for all but one expert endoscopist.

► **Table 3** Overview of prospective clinical trials assessing performance of deep learning-based computer-aided diagnosis systems for endoscopic differentiation between neoplastic and non-neoplastic polyps.

| | Year | Multi-center | Classification groups | Imaging modality | Polyps assessed for clinical validation | Endoscopist comparison group (experience) | Results | | | | | | | | | |
|------------------------------|------|--------------|---|------------------------|--|---|---|--|---|-----|----------------|-----|-----|-----|-----|--|
| | | | | | | | CADx† | | Endoscopists | | CADx-assisted† | | | | | |
| | | | | | | | All | RS | All | RS | All | RS | | | | |
| Kominami et al. [77] | 2016 | No | Adenomas vs. non-adenomas (SSLs excluded) | Magnified NBI | 88 diminutive polyps (45 AD, 43 NAD) of which 54 in RS | N/A | Success rate CADx (%) HC diagnosis (%) Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | N/A N/A 88 93.2 93.0 93.3 93.0 93.3 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | |
| Mori et al. [78]§ | 2018 | No | Adenomas vs. non-adenomas (SSLs excluded) | Endocytoscopy with NBI | 466 diminutive polyps (287 AD, 175 NAD) of which 250 in RS | N/A† | Success rate CADx (%) HC diagnosis (%) Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | 98.1 N/A 466 91.6 92.7 89.8 93.7 88.3 | N/A N/A 250 94.4 93.3 95.2 93.3 95.2 | N/A | N/A | N/A | N/A | N/A | N/A | |
| Barua et al. [79] | 2022 | Yes | Neoplastic (including SSLs for primary analyses) vs. non-neoplastic (including SSLs for secondary analyses) | (Ultra)-magnified NBI | 892 diminutive polyps (359 NP, 533 NNP), all in RS | 22 endoscopists (all 1–5 years of colonoscopy experience or performed 200–1000 colonoscopies) | Success rate CADx (%) HC diagnosis (%)†† Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | N/A N/A 892 88.4 83.1 78.9 91.5 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | 92.6 892 N/A 90.4 85.9 82.0 92.8 |
| García-Rodríguez et al. [80] | 2022 | No | Adenomas vs. non-adenomas (including SSLs) | WLE | 52 diminutive polyps (35 AD, 18 NAD) of which unspecified amount in RS | 4 endoscopists (all staff endoscopists without specific training in optical diagnosis) | Success rate CADx (%) HC diagnosis (%)‡ Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | N/A N/A 52 78.8 88.2 61.1 81.1 73.3 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | 52 71.1 58.5* 94.4* 95.2 54.8 |

▶ **Table 3** (Continuation)

| | Year | Multi-center | Classification groups | Imaging modality | Polyps assessed for clinical validation | Endoscopist comparison group (experience) | Success rate CADx (%) | Results | | | | | |
|------------------------|------|--------------|--|------------------|---|---|---|---|--|--|--|---|----|
| | | | | | | | | CADx† | | Endoscopists | | CADx-assisted† | |
| | | | | | | | | All | RS | All | RS | All | RS |
| Hassan et al. [81]§§ | 2022 | No | Adenomas vs. non-adenomas (including SSLs) | WLE§§ | 476 diminutive polyps (163 AD, 291 NAD) of which 295 in RS | 4 endoscopists (all > 2000 screening colonoscopies, training in optical diagnosis) | Success rate CADx (%) HC diagnosis (%) Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | 95.4 N/A 454 86.8 82.0 89.5 81.0 90.1 | 98.6 N/A 291 91.8 82.0 93.2 65.3 97.6 | N/A N/A N/A | N/A 92.2 439 90.7 80.6 95.5 89.2 91.3 | N/A 94.6 279 96.1 81.2 98.0 83.9 97.6 | |
| Houwen et al. [82] | 2023 | Yes | Neoplastic (including SSLs) vs. non-neoplastic | NBI | 429 diminutive polyps (300 AD, 41 SSLs, 82 HPs) of which 122 in RS | 20 endoscopists (all BCSP certified) | Success rate CADx (%) HC diagnosis (%)†† Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | 98.3 99.7 422 79.6 89.4 38.3 85.9 46.3 | N/A 99.2 121 75.2 91.7 37.8 77.0 66.7 | N/A 86.8 367 85.8 94.7 47.1 88.7 66.7 | N/A 98.1 415 84.1 93.2 44.9 88.8 60.3 | N/A 95.9 117 84.6 90.2 71.4* 88.1 75.8 | |
| Minegishi et al. [83] | 2022 | No | Neoplastic (including SSLs) vs. non-neoplastic | NBI | 395 diminutive polyps (259 AD, 25 SSL, 11 NNP) of which at least 126 in RS (exact amount unspecified) | 16 endoscopists (11 board certified experts with ≥ 5 years experience, 5 non-experts) | Success rate CADx (%) HC diagnosis (%) Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | 94.2 N/A 372 84.4 93.3 61.5 N/A N/A | N/A N/A 372 85.5 94.4 62.5 N/A N/A | N/A 92.7 366 88.0 95.8 67.0 88.5 85.9 | N/A N/A 126 85.7 95.2 76.2 80.0 94.1 | | |
| Rondonotti et al. [84] | 2022 | Yes | Adenomas vs. non-adenomas (including SSLs) | BLI | 596 diminutive polyps (259 AD, 337 NAD), all in RS | 18 endoscopists (9 expert endoscopists, 9 non-expert endoscopists†††) | Success rate CADx (%) HC diagnosis (%) Number of polyps Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | N/A N/A 541 85.8 81.9 88.7 84.4 86.7 | 90.8 N/A 541 85.8 81.9 88.7 84.4 86.7 | N/A N/A 540 88.7 88.6 88.8 86.1 90.1 | N/A 92.3 550 88.4 88.6 88.1 85.1 91.0 | | |

▲ **Table 3** (Continuation)

CADx, computer-aided diagnosis; SSL, sessile serrated lesion; NBI, narrow band imaging; AD, adenomas; NAD, non-adenomas; RS, rectosigmoid; N/A, not available; HC, high confidence; PPV, positive predictive value; NPV, negative predictive value; NP, neoplastic polyp; NNP, non-neoplastic polyp; WLE, white light endoscopy; HP, hyperplastic polyp; BCSP, bowel cancer screening program; BLI, blue light imaging.

* Significant difference compared to CADx performance ($P < 0.05$).

† If available, performance with inclusion of only high-confidence diagnoses reported.

‡ Only concordance between optical diagnosis by endoscopists and CADx system diagnosis reported (97.5%).

§ Reported results concern 'worst-case scenario' results as reported in study (9 polyps for which CADx system diagnosis was not possible were treated as either false-positive or false-negative).

¶ CADx performance benchmarked to endoscopists in separate (non-real-time) test set, hence not reported within table.

‡‡ High-confidence diagnosis cut-off threshold: 70%.

‡‡‡ High-confidence diagnosis cut-off threshold: 80%.

§§ Within this study, CADx trained and tested for white light endoscopy. However, endoscopists used virtual chromoendoscopy for optical diagnosis and were not blinded for the CADx diagnosis. While unblinded to CADx diagnosis, endoscopist performance is reported as CADx-assisted performance.

¶¶ High-confidence diagnosis cut-off threshold: 50%.

¶¶¶ Experts endoscopists followed a dedicated training program, underwent periodic auditing and monitoring and performed optical diagnosis on a regular basis. Non-experts were endoscopists that did not fulfill these criteria.

[11, 12, 13]. Despite these facts, computer-aided diagnosis of polyp morphology is a field yet to be explored: to the best of our knowledge, only one study describing assessment of polyp morphology by a computer system, as part of an algorithm for automated textual polyp image description, is available [75].

Computer-aided prediction of polyp histology

Differentiation of diminutive neoplastic from non-neoplastic (or adenomatous from non-adenomatous) polyps

Colorectal polyps can generally be subdivided into neoplastic and non-neoplastic. Neoplastic lesions concerns both lesions yielding malignant potential and malignant lesions, while non-neoplastic lesions do not yield malignant potential. Hence, removal and analysis of non-neoplastic lesions is often unnecessary [23, 24]. Real-time optical differentiation of neoplastic and non-neoplastic polyps during colonoscopy procedures could help to reduce the significant patient and economic burden caused by unnecessary resection and analysis of non-neoplastic lesions [76]. For this reason, the 'leave-in-situ' and 'resect-and-discard' optical diagnosis strategies have been proposed [23, 24]. However, while proposed optical diagnosis performance thresholds are frequently not met in community practice, feasibility of these strategies is still limited [14, 15, 16].

A wide variety of studies describing computer systems trained to differentiate neoplastic and non-neoplastic lesions based on polyp phenotype has been published. For this review, we will highlight available prospective clinical trials evaluating the performance of such systems in real-time clinical settings and using either white light, (magnified) narrow band imaging (NBI) or blue light imaging (BLI) imaging modalities (► **Table 3**) [77, 78, 79, 80, 81, 82, 83, 84]. In these studies, overall accuracies of computer-aided diagnosis (CADx) systems ranged between 78.8% and 93.2% [77, 78, 80, 81, 82, 83]. Reported accuracies for diminutive polyps located within the rectosigmoid ranged between 75.2% and 94.4% [78, 81, 82, 84]. Within five studies, CADx system performance was benchmarked to performance by endoscopists [80, 81, 82, 83, 84]. Two studies reported significant differences, both in favor of the endoscopists and the CADx system [80, 82]. Four studies reported performance of endoscopists with real-time assistance of a CADx system [79, 81, 82, 83, 84]. While no significant benefits were reported for computer-aided colonoscopy when compared to endoscopists alone, one of these studies did show that non-experts can eventually meet expert accuracy levels when performing real-time computer-aided polyp assessment on a regular basis [84].

To facilitate implementation of optical diagnosis strategies in daily practice, the Preservation and Incorporation of Valuable Endoscopic Innovations (PIVI) initiative [23] and Simple Optical Diagnosis Accuracy (SODA) [24] competence standards have been described. In ► **Table 4**, results of described clinical trials are evaluated along the lines of these standards. While most

Table 4 Performance of deep learning-based computer-aided diagnosis systems for endoscopic differentiation of neoplastic from non-neoplastic polyps evaluated in the context of PIVI and SODA criteria.

| Threshold definition | | Kominami et al. [77] ^a | Mori et al. [78] [*] | Barua et al. [79] | Garcia-Rodríguez et al. [80] ^{†‡} | Hassan et al. [81] [*] | Houwen et al. [82] [§] | Minegishi et al. [83] [*] | Rondonotti et al. [84] [§] |
|----------------------|---|-----------------------------------|-------------------------------|-------------------|--|---------------------------------|---------------------------------|------------------------------------|--|
| PIVI-1 | ≥ 90% agreement with USMSTF or ESCE post-polypectomy guidelines | N/A [¶] N/A | N/A N/A | N/A N/A | N/A N/A | 95.9 (yes) 95.6 (yes) | N/A 95.5 (yes) | 90.1 (yes) 93.4 (yes) | 92.1 (yes) 96.8 (yes) |
| PIVI-2 | ≥ 90% negative predictive value for neoplastic lesions in the rectosigmoid | N/A ^{**} | 95.2 (yes) | N/A | N/A | 97.6 (yes) | 78.6 (no) | N/A | 86.7 (no) |
| SODA-1 | ≥ 90% sensitivity and 80% specificity for high-confidence endoscopic characterization of colorectal neoplasia of 1–5 mm in the rectosigmoid | N/A N/A | 93.3 (yes) 95.2 (yes) | N/A N/A | N/A N/A | 82.0 (no) 93.2 (yes) | 91.7 (yes) 37.8 (no) | N/A N/A | 81.9 (no) 88.7 (yes) |
| SODA-2 | ≥ 80% sensitivity and 80% specificity for high-confidence endoscopic characterization of colorectal neoplasia of 1–5 mm | 93.0 (yes) 93.3 (yes) | 92.7 (yes) 89.8 (yes) | N/A N/A | 88.2 (yes) 61.1 (no) | 82.0 (yes) 89.5 (yes) | 89.4 (yes) 38.3 (no) | 93.3 (yes) 61.5 (no) | 81.9 (yes) ^{††} 88.7 (yes) ^{††} |

PIVI, Preservation and Incorporation of Valuable Endoscopic Innovations; SODA, Simple Optical Diagnosis Accuracy; USMSTF, United States Preventive Services Taskforce; ESCE, European Society of Gastrointestinal Endoscopy; N/A, not available; NPV, negative predictive value.

^{*}No specific information concerning differentiation between low- and high-confidence diagnoses by the CADx system reported within this study. Reported results based on all included polyps or all polyps for which a CADx system diagnosis was available.

[†]Stand-alone CADx performance not reported within this study.

^{†‡}Differentiation between low- and high-confidence diagnoses by the CADx system reported within this study (threshold cut-off value reported in **Table 3**), but results for high-confidence CADx system diagnoses only not available. Reported results based on both low- and high-confidence CADx system diagnoses.

[§]Differentiation between low- and high-confidence diagnoses by the CADx system reported within this study (threshold cut-off value reported in **Table 3**). Reported results based on high-confidence CADx system diagnoses only.

[¶]No guideline agreement reported for diminutive polyps specifically (guideline agreement for all included polyps: 92.7%).

^{**}No NPV reported for diminutive rectosigmoid polyps specifically (NPV for all detected diminutive polyps: 93.3%).

^{††}Study only includes rectosigmoid polyps, hence similar to performance as reported for SODA-1 criterion.

CADx systems were able to meet several of the performance thresholds, none of them met all thresholds. This does however also relate to the fact that only two studies reported all required parameters [82,84]. Besides, an important issue to address is that, according to PIVI and SODA standards, only high-confidence (HC) diagnoses should be used to calculate performance parameters [23,24]. Nonetheless, within most studies differentiation between high- and low-confidence CADx system diagnoses was either not described [77,78,81,84], or results with and without inclusion of low-confidence diagnoses were not separately reported [79,80,83]. Moreover, a standard HC diagnosis threshold cut-off for CADx systems is lacking (i. e. threshold concerning the minimum degree of certainty that an algorithm requires to consider an output a HC diagnosis). This results in CADx systems adopting different HC diagnosis threshold cut-offs [79,80,82], making reliable comparison and evaluation impossible.

From clinical perspective, the fact that different studies managed sessile serrated lesions (SSLs) in different ways should also be addressed. While SSLs are estimated to make up 15–30% of CRC cases [85] and especially optical differentiation between SSLs (neoplastic) and hyperplastic polyps (non-neoplastic) is known to be challenging [86], only two studies used a CADx system that was specifically trained for recognition of SSLs [82,83]. Besides, only three studies (partly) included SSLs within the neoplastic polyp group [79,82,83], while others assigned SSLs to the non-neoplastic group [80,81,84] or excluded all SSLs [77,78]. Additional limitations relate to the fact that the number of included polyps was low in several studies, most studies were single center and only two studies involved ‘non-expert’ endoscopists [79,84].

Despite remaining limitations and need for further optimization of system performances to reach PIVI and SODA thresholds, most CADx systems for differentiation of neoplastic and non-neoplastic lesions showed to be able to meet expert endoscopist performance in real-time clinical settings. In addition, a significant optical diagnosis learning curve for ‘non-expert’ endoscopists was illustrated [84]. In the last place, CADx showed the potential to increase the proportion of HC diagnoses by endoscopists compared to unaided optical diagnosis [79,82]. This is crucial to establish a reduction in unnecessary polypectomies and pathological assessments [23,24]. Hence, with the commercial availability of some of the evaluated CADx systems [87], first steps toward improved polyp assessment with use of CADx systems might be ahead.

Differentiation between polyps with different degrees of dysplasia

Several studies assessing the feasibility of deep learning approaches for differentiation of polyps with different degrees of dysplasia (LGD versus HGD) are available (Table S1) [88,89,90,91,92,93]. This is relevant as lesions harboring HGD should ideally be resected en bloc [4,25] and may warrant shortened surveillance intervals [27,28].

Reported accuracies in six identified studies ranged between 80.2 and 94.6% [88,89,90,91,92,93]. In three of these studies, the CADx systems outperformed endoscopists with different

levels of experience [88,91,93]. However, none of the proposed systems was evaluated in a real-time clinical setting and most studies also included lesions other than lesions with LGD or HGD.

Because the prevalence of HGD in diminutive polyps is low [94,95,96], the additional value of these systems for optical diagnosis strategies is uncertain. Nonetheless, they may be useful for development of algorithms for purposes such as identification of areas with advanced dysplasia in larger lesions. Moreover, while most algorithms are also trained for recognition of adenocarcinoma, these algorithms might be useful to address clinical challenges such as endoscopic recognition of T1 CRCs [97,98].

Differentiation between superficial and deep invasive lesions

In case of a suspected CRC, the choice and feasibility of en bloc resection methods depends on the depth of invasion [25,26]. Nonetheless, imaging modalities to accurately determine lesion invasion depth are lacking. Hence, differentiation of lesions with and without DSI is mostly done based on endoscopic identification of specific morphological polyp features [20,21,71,72] and surface characteristics [99,100] that are known to be associated with DSI. However, this endoscopic differentiation is known to be challenging [20,21,22].

Deep learning approaches for differentiation of lesions with and without DSI have been proposed in several studies (Table 5) [101,102,103,104,105,106,107,108]. Identified studies reported accuracies ranging between 81.2% and 94.1% [101,102,103,104,105,106,107,108]. Some of these studies benchmarked CADx system performance to performance of endoscopists with variable degrees of experience [102,104,105,106,107,108]. In a few studies, the CADx system outperformed one or more of the novices and trainees [104,105,106,108]. In addition, one study illustrated that diagnostic accuracy of endoscopists improved with assistance of a CADx system [108]. However, in none of the studies the CADx system was able to significantly outperform experienced or expert endoscopists.

Although these results seem promising, they should be carefully interpreted. Firstly, none of the systems was validated in a real-time clinical setting. Moreover, CADx systems were trained and validated using different imaging modalities, with two studies showing that performance may differ per imaging modality [104,106]. Besides, the datasets considerably differed in both size and composition. Only three studies reported CADx systems that were tested on datasets consisting of CRCs only (both with and without DSI) [103,105,106] while other studies also included benign lesions in the non-DSI group [101,102,104,107,108].

With recent introduction of new endoscopic resection methods, possibilities for local resection for lesions with DSI seem to be increasing. As a result, it could be debated whether optical diagnosis should not be adapted to also differentiate lesions with different degrees of DSI [109]. This might also have implications for future development of CADx systems designed for

► **Table 5** Overview of studies describing deep learning approaches for endoscopic differentiation of lesions with and without deep submucosal invasion.

| Year | Included lesions within test set | Imaging modality | Datasets† | Endoscopist comparison group(s) (experience) | Results | Endoscopists | | | | | | |
|-----------------------|----------------------------------|---|--|--|---|---|--|--------------------------------------|--------------------------------------|--------------------------------------|-----|-----|
| | | | | | | CADx | G1 | G2 | G3 | | | |
| Takeda et al. [101] | 2017 | Non-invasive (AD) vs. invasive (CRC with DSI) | Endocytoscopy | Training set: 5543 images Test set: 200 images | N/A | 94.1 [†] 89.4 [‡] 98.9 [†] 98.8 [‡] 90.1 [†] | Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | N/A | N/A | N/A | | |
| Tamai et al. [102] | 2017 | Non-invasive (HPS, AD, CRC without DSI) vs. invasive (CRC with DSI) | Magnified NBI | Training set: N/A Test set: 121 images | Group 1: Two experienced endoscopists (experience undefined) | 88.4 55.0 95.0 68.8 91.4 | Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | 91.7 70.0 96.1 77.8 94.2 | N/A | N/A | N/A | |
| Ito et al. [103] | 2019 | Non-invasive (CRC without DSI) vs. invasive (CRC with DSI) | WLE | Training set: 9,942 images Test set: 5,022 images | N/A | 81.2 67.5 89.0 N/A N/A | Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | N/A | N/A | N/A | N/A | |
| Lui et al. [104] | 2019 | Non-invasive (SSL, AD, CRC without DSI) vs. invasive (CRC with DSI) | WLE, NBI | Training set: 8000 images Test set: 567 images | Group 1: One expert (> 2000 colonoscopies) Group 2: Two trainees (> 500 colonoscopies, NBI training) | 85.5 [†] 88.2 [†] 77.9 [†] 92.1 [†] 69.3 [†] | Accuracy (%) [§] Sensitivity (%) [§] Specificity (%) [§] PPV (%) [§] NPV (%) [§] | 86.4 91.8 52.6 92.4 50.6 | 71.0 73.9 61.8 85.7 47.2 | N/A | N/A | |
| Nakajima et al. [105] | 2020 | Non-invasive (CRC without DSI) vs. invasive (CRC with DSI) | WLE | Training set: 1839 images Test set: 78 images | Group 1: Two experts (> 5000 colonoscopies) Group 2: Two trainees (< 500 colonoscopies) Group 3: Two novices (15 minutes education by case study, < 6 months residency) | 84.1 81.0 87.0 85.0 83.3 | Accuracy (%) ^{**} Sensitivity (%) ^{**} Specificity (%) ^{**} PPV (%) ^{**} NPV (%) ^{**} | 92.0 85.7 97.8 97.3 88.2 | 71.5 83.3 60.9 65.4 83.3 | 65.9 100 34.8 58.7 100 | | |
| Lu et al. [106] | 2021 | Non-invasive (CRC without DSI) vs. invasive (CRC with DSI) | WLE, magnified and non-magnified NBI/BLI | Training set: 21,433 images Test set: 168 images | Group 1: Three experts (> 10 years' experience) Group 2: Five seniors (> 5 years' experience) Group 3: Seven juniors (> 1 year experience) | 88.1 92.5 85.6 79.0 95.0 | Accuracy (%) ^{††} Sensitivity (%) ^{††} Specificity (%) ^{††} PPV (%) ^{††} NPV (%) ^{††} | 84.1 64.2 95.8 91.4 82.1 | 84.1 64.2 95.8 91.4 82.1 | 76.7 62.3 85.4 76.1 80.3 | | |
| Luo et al. [107] | 2021 | Non-invasive (AD, CRC without DSI) vs. invasive (CRC with DSI) | WLE | Training set: 7734 images Test set: 1634 images | Group 1: Two experts (> 200 ESD cases), six experienced endoscopists (> 3000 colonoscopies, > 200 EMR cases, > 30 ESD cases) | 91.1 91.2 91.0 86.7 93.7 | Accuracy (%) Sensitivity (%) Specificity (%) PPV (%) NPV (%) | 92.6 88.4 95.5 93.2 92.2 | N/A | N/A | N/A | N/A |

▶ **Table 5** (Continuation)

| | Year | Included lesions within test set | Imaging modality | Datasets† | Endoscopist comparison group(s) (experience) | Results | | |
|-----------------------|------|---|------------------|---|--|--------------------------------------|---|-----|
| | | | | | | CADx | Endoscopists | |
| Tokunaga et al. [108] | 2021 | Non-invasive (AD, CRC without DSI) vs. invasive (CRC without DSI) | WLE | Training set: 2751 images Test set: 691 images | Group 1: Two experts (> 2000 colonoscopies + 500 EMRs/ESDs) Group 2: Two trainees (< 500 colonoscopies) | 90.3 96.7 75.0 90.2 90.5 | 89.4 96.5 72.5 89.4 89.7 84.9* 92.1* 67.6* 87.2 78.2 | N/A |

CADx, computer-aided diagnosis; G1, group 1; G2, group 2; G3, group 3; AD, adenoma; CRC, colorectal cancer; DSI, deep submucosal invasion; N/A, not available; PPV, positive predictive value; NPV, negative predictive value; HP, hyperplastic polyp; NBI, narrow band imaging; WLE, white light endoscopy; SSL, sessile serrated lesion; BLI, blue light imaging; EMRs, endoscopic mucosal resections; ESDs, endoscopic submucosal dissections.

*Significant difference compared to CADx performance ($P < 0.05$).

†Data used for internal validation is reported as part of the training set.

‡Reported results concern results including both high- and low-confidence diagnoses. Performance for high-confidence diagnoses only (72.0% of included polyps): accuracy 99.3%, sensitivity 98.1%, specificity 100%, PPV 100%, NPV 98.8%.

§Reported results for endoscopists in group 2 (trainees) concern calculated means. No P value for comparison of mean to CADx system performance available. On individual basis, the CADx system significantly outperformed one of the trainee endoscopists on accuracy, sensitivity, and NPV. The other trainee endoscopist was significantly outperformed by the CADx system on accuracy, specificity, and PPV.

¶Reported results concern average performance of CADx system for both narrow band imaging and white light endoscopy images.

**Reported results for endoscopists concern calculated means for separate endoscopist groups. No P value for comparison of means to CADx system performance available. On individual basis, the CADx system reached significantly higher accuracy than one of the trainee endoscopists and one of the novice endoscopists. CADx system reached significantly higher specificity than both novice endoscopists and one trainee endoscopist reached significantly higher sensitivity than the CADx system.

††Reported results for endoscopists concern calculated means for separate endoscopist groups. No P value for comparison of means to CADx system performance available. On individual basis, CADx system significantly outperformed five of seven junior endoscopists on overall performance, but none of the senior or expert endoscopists.

‡‡No significance levels for difference between endoscopist and CADx system performance available for NPV and PPV.

assessment of CRC invasion depth. However, clinical validation of currently available CADx systems is warranted first.

Discussion

Over the past decade, advances in AI and computer science have led to an exponential increase in studies on computer-aided diagnosis of colorectal polyps. As outlined within this review, the most substantial developments in the field of computer-aided polyp diagnosis involve CADx systems for differentiation between neoplastic and non-neoplastic lesions. Several studies have demonstrated potential of such systems to meet expert performance levels in real-time clinical settings. Developmental processes of computer systems that are able to provide real-time feedback to endoscopists on polyp characteristics such as size, location, degree of dysplasia and invasion depth are still in preliminary phases. Future studies should mainly focus on prospective clinical validation of these systems. Besides, feasibility of CADx systems for specific assessment of polyp morphology has yet to be explored.

Adopting computer systems for colorectal polyp assessment in daily practice might yield several benefits. Primarily, if these systems are trained with high quality expert-annotated data, they could possibly serve as an objective, expert-level second observer that is not prone to human factors such as fatigue, distraction or subjectivity. Especially for less experienced endoscopists, this could provide opportunities to optimize accuracy of polyp assessments, thereby possibly improving clinical outcomes and reducing patient burden and costs. In addition, availability of computer systems able to assess independent polyp characteristics could provide possibilities for automated polyp description for endoscopy reports [75]. When combined with algorithms for purposes such as recognition of resection methods [110], this might significantly ease administrative burdens for endoscopists. In the last place, optimizing accuracy of endoscopic assessment of different polyp characteristics could aid in development of more trustworthy clinical decision-making algorithms or prediction models involving specific polyp characteristics [111, 112, 113].

On the other hand, clinicians should also be aware of the limitations and potential disadvantages of computer-aided polyp diagnosis. Especially systems based on machine learning architectures are highly dependent on the training data used. While these systems are often trained with human-annotated data, these systems are not likely to outperform experts on a regular basis. Therefore, clinicians should be aware that these systems are not flawless. In addition, system performance is also dependent on what is shown by the endoscopist: the quality of the images provided to the computer system during endoscopies might differ between endoscopists, possibly influencing system performance and feasibility [82]. Moreover, it can be hypothesized that regular CADx system-assisted colonoscopy might eventually lead to a certain degree of user-dependency.

There are also several more general issues to be addressed when considering the future perspectives of CADx systems in endoscopy practice. In the first place, insights into the cost-effectiveness of CADx systems are still scarce. Although it is sug-

gested that CADx could potentially lead to a 11% reduction of average colonoscopy costs [114], figures concerning actual cost reduction due to use of CADx systems in different countries and clinical settings are still lacking. Second, there might be limitations concerning the technical integration of CADx systems in different endoscopy suites and settings, while most systems have unique hardware and software requirements and are not simply compatible with all regularly used endoscopy devices. Third, the sentiment of physicians toward AI and computer-aided diagnosis should be taken in consideration: increased costs, operator dependency and increased procedural time are common concerns among physicians [115]. Moreover, basic technical knowledge on topics such as machine learning is warranted to be able to critically appraise available literature on the topic of computer-aided diagnosis approaches and appraise the possible technical biases inherent to available systems. Due to the novelty of AI and computer-aided diagnosis, most clinicians will likely lack this knowledge. Therefore, specific education and training will be needed to increase its feasibility.

Despite the various limitations and uncertainties, it should be emphasized that computer-aided diagnosis has only been a topic of interest within the field of gastrointestinal endoscopy for a little over ten years. Hence, especially in the context of the rapidly increasing amounts of research on this topic, toward the future computer-aided diagnosis will likely take a more prominent role in daily endoscopy practice. On one hand this relates to the fact that (technical) innovations in upcoming years will likely aid in improving accuracy of existing CADx systems, while there are also still numerous purposes for which possibilities of computer-aided diagnosis is yet to be explored. In example, besides computer systems that could aid endoscopists in assessment of polyp morphology, systems for purposes such as suggestion of appropriate polyp resection method or assessment of completeness of resection might yield significant clinical potential.

The strength of this review is that, to the best of our knowledge, this is the first review to provide such a broad overview of available studies on computer-aided diagnosis of all polyp characteristics essential for clinical decision making. However, in the context of the extensive scope of the aim of this review, we decided to comply to a narrative rather than a systematic review approach. While this might have resulted in accidental miss of relevant publications, this can be considered a limitation.

Conclusions

To conclude, with recent breakthroughs in the field of AI and computer science, a major increase in research on the topic of computer-aided colorectal polyp assessment is seen. With commercial availability of CADx systems for differentiation between neoplastic and non-neoplastic polyps, first steps toward improved endoscopy quality with use of CADx systems in daily practice might be ahead. However, optimization of performance is still required to ensure that these CADx systems meet all performance thresholds. Besides, toward the future, further innovation, exploration and clinical validation of computer-ai-

ded diagnosis approaches for diagnosis of other polyp characteristics is required for realization of complete computer-aided polyp assessment.

Conflict of Interest

ED received a research grant from Fujifilm, a consulting fee for medical advice from Olympus, Fujifilm, GI Supply, PAION, Ambu and CPP-FAP and a speakers' fee from Olympus, Roche, GI Supply, Norgine, Fujifilm and IPSEN. PF received research support from Boston Scientific and a consulting fee from Olympus and Cook Endoscopy. The remaining Authors declare that there is no conflict of interest.

References

- [1] Sung H, Ferlay J, Siegel RL et al. Global Cancer Statistics 2020: GLOBOCAN estimates of incidence and mortality worldwide for 36 cancers in 185 countries. *CA Cancer J Clin* 2021; 71: 209–249 doi:10.3322/caac.21660
- [2] Nguyen LH, Goel A, Chung DC. Pathways of colorectal carcinogenesis. *Gastroenterology* 2020; 158: 291–302 doi:10.1053/j.gastro.2019.08.059
- [3] Rex DK, Boland CR, Dominitz JA et al. Colorectal cancer screening: Recommendations for physicians and patients from the U.S. Multi-Society Task Force on Colorectal Cancer. *Gastrointest Endosc* 2017; 86: 18–33
- [4] Saftoiu A, Hassan C, Areia M et al. Role of gastrointestinal endoscopy in the screening of digestive tract cancers in Europe: European Society of Gastrointestinal Endoscopy (ESGE) Position Statement. *Endoscopy* 2020; 52: 293–304 doi:10.1055/a-1104-5245
- [5] Zauber AG. The impact of screening on colorectal cancer mortality and incidence: has it really made a difference? *Dig Dis Sci* 2015; 60: 681–691 doi:10.1007/s10620-015-3600-5
- [6] O'Connor SA, Hewett DG, Watson MO et al. Accuracy of polyp localization at colonoscopy. *Endosc Int Open* 2016; 4: E642–E646 doi:10.1055/s-0042-105864
- [7] de Vries AH, Bipat S, Dekker E et al. Polyp measurement based on CT colonography and colonoscopy: variability and systematic differences. *Eur Radiol* 2010; 20: 1404–1413 doi:10.1007/s00330-009-1683-0
- [8] Moug SJ, Vernall N, Saldanha J et al. Endoscopists' estimation of size should not determine surveillance of colonic polyps. *Colorectal Dis* 2010; 12: 646–650
- [9] Eichenseer PJ, Dhanekula R, Jakate S et al. Endoscopic mis-sizing of polyps changes colorectal cancer surveillance recommendations. *Dis Colon Rectum* 2013; 56: 315–321 doi:10.1097/DCR.0b013e31826dd138
- [10] Chaptini L, Chaaya A, Depalma F et al. Variation in polyp size estimation among endoscopists and impact on surveillance intervals. *Gastrointest Endosc* 2014; 80: 652–659
- [11] van Doorn SC, Hazewinkel Y, East JE et al. Polyp morphology: an interobserver evaluation for the Paris classification among international experts. *Am J Gastroenterol* 2015; 110: 180–187
- [12] Kim JH, Nam KS, Kwon HJ et al. Assessment of colon polyp morphology: Is education effective? *World J Gastroenterol* 2017; 23: 6281–6286 doi:10.3748/wjg.v23.i34.6281
- [13] Lee YJ, Kim ES, Park KS et al. Interobserver agreement in the endoscopic classification of colorectal laterally spreading tumors: a multicenter study between experts and trainees. *Dig Dis Sci* 2014; 59: 2550–2556
- [14] Kuiper T, Marsman WA, Jansen JM et al. Accuracy for optical diagnosis of small colorectal polyps in nonacademic settings. *Clin Gastroenterol Hepatol* 2012; 10: 1016–1020 ; quiz e1079
- [15] Ladabaum U, Fioritto A, Mitani A et al. Real-time optical biopsy of colon polyps with narrow band imaging in community practice does not yet meet key thresholds for clinical decisions. *Gastroenterology* 2013; 144: 81–91
- [16] Patel SG, Schoenfeld P, Kim HM et al. Real-time characterization of diminutive colorectal polyp histology using narrow-band imaging: implications for the resect and discard strategy. *Gastroenterology* 2016; 150: 406–418 doi:10.1053/j.gastro.2015.10.042
- [17] Ignjatovic A, East JE, Suzuki N et al. Optical diagnosis of small colorectal polyps at routine colonoscopy (Detect InSpect Characterise Resect and Discard; DISCARD trial): a prospective cohort study. *Lancet Oncology* 2009; 10: 1171–1178 doi:10.1016/S1470-2045(09)70329-8
- [18] Komeda Y, Kashida H, Sakurai T et al. Magnifying narrow band imaging (nbi) for the diagnosis of localized colorectal lesions using the Japan NBI Expert Team (JNET) classification. *Oncology* 2017; 93: 49–54 doi:10.1159/000481230
- [19] Sumimoto K, Tanaka S, Shigita K et al. Diagnostic performance of Japan NBI Expert Team classification for differentiation among non-invasive, superficially invasive, and deeply invasive colorectal neoplasia. *Gastrointest Endosc* 2017; 86: 700–709
- [20] Saitoh Y, Obara T, Watari J et al. Invasion depth diagnosis of depressed type early colorectal cancers by combined use of videoendoscopy and chromoendoscopy. *Gastrointest Endosc* 1998; 48: 362–370 doi:10.1016/s0016-5107(98)70004-5
- [21] Horie H, Togashi K, Kawamura YJ et al. Colonoscopic stigmata of 1 mm or deeper submucosal invasion in colorectal cancer. *Dis Colon Rectum* 2008; 51: 1529–1534 doi:10.1007/s10350-008-9263-y
- [22] Backes Y, Schwartz MP, Ter Borg F et al. Multicentre prospective evaluation of real-time optical diagnosis of T1 colorectal cancer in large non-pedunculated colorectal polyps using narrow band imaging (the OPTICAL study). *Gut* 2019; 68: 271–279 doi:10.1136/gutjnl-2017-314723
- [23] Rex DK, Kahi C, O'Brien M et al. The American Society for Gastrointestinal Endoscopy PIVI (Preservation and Incorporation of Valuable Endoscopic Innovations) on real-time endoscopic assessment of the histology of diminutive colorectal polyps. *Gastrointest Endosc* 2011; 73: 419–422 doi:10.1016/j.gie.2011.01.023
- [24] Houwen B, Hassan C, Coupé VMH et al. Definition of competence standards for optical diagnosis of diminutive colorectal polyps: European Society of Gastrointestinal Endoscopy (ESGE) Position Statement. *Endoscopy* 2022; 54: 88–99 doi:10.1055/a-1689-5130
- [25] Ferlitsch M, Moss A, Hassan C et al. Colorectal polypectomy and endoscopic mucosal resection (EMR): European Society of Gastrointestinal Endoscopy (ESGE) Clinical Guideline. *Endoscopy* 2017; 49: 270–297 doi:10.1055/s-0043-102569
- [26] Kaltenbach T, Anderson JC, Burke CA et al. Endoscopic removal of colorectal lesions: Recommendations by the US Multi-Society Task Force on Colorectal Cancer. *Am J Gastroenterol* 2020; 115: 435–464 doi:10.14309/ajg.0000000000000555
- [27] Hassan C, Antonelli G, Dumonceau JM et al. Post-polypectomy colonoscopy surveillance: European Society of Gastrointestinal Endoscopy (ESGE) Guideline – Update 2020. *Endoscopy* 2020; 52: 687–700 doi:10.1055/a-1185-3109
- [28] Gupta S, Lieberman D, Anderson JC et al. Recommendations for follow-up after colonoscopy and polypectomy: A consensus update by the US Multi-Society Task Force on Colorectal Cancer. *Am J Gastroenterol* 2020; 115: 415–434 doi:10.14309/ajg.0000000000000544
- [29] Chowdhury M, Sadek AW. Advantages and limitations of artificial intelligence. *Artificial Intelligence Applications to Critical Transportation Issues. Transportation Research Circular E-C168.*

- [30] van der Sommen F, de Groof J, Struyvenberg M et al. Machine learning in GI endoscopy: practical guidance in how to interpret a novel field. *Gut* 2020; 69: 2035–2045 doi:10.1136/gutjnl-2019-320466
- [31] Fujita H. AI-based computer-aided diagnosis (AI-CAD): the latest review to read first. *Radiol Phys Technol* 2020; 13: 6–19 doi:10.1007/s12194-019-00552-4
- [32] Hassan C, Spadaccini M, Iannone A et al. Performance of artificial intelligence in colonoscopy for adenoma and polyp detection: a systematic review and meta-analysis. *Gastrointest Endosc* 2021; 93: 77–85 e76 doi:10.1016/j.gie.2020.06.059
- [33] Deliwala SS, Hamid K, Barbarawi M et al. Artificial intelligence (AI) real-time detection vs. routine colonoscopy for colorectal neoplasia: a meta-analysis and trial sequential analysis. *Int J Colorectal Dis* 2021; 36: 2291–2303 doi:10.1007/s00384-021-03929-3
- [34] Segev L, Kalady MF, Plessec T et al. The location of premalignant colorectal polyps under age 50: a further rationale for screening sigmoidoscopy. *Int J Colorectal Dis* 2020; 35: 529–535
- [35] Stauffer CMP. *Colonoscopy*. Treasure Island (FL): StatPearls Publishing; 2022
- [36] Phillips M, Patel A, Meredith P et al. Segmental colonic length and mobility. *Ann R Coll Surg Engl* 2015; 97: 439–444 doi:10.1308/003588415X14181254790527
- [37] Saunders BP, Phillips RK, Williams CB. Intraoperative measurement of colonic anatomy and attachments with relevance to colonoscopy. *Br J Surg* 1995; 82: 1491–1493 doi:10.1002/bjs.1800821113
- [38] Madiba TE, Haffajee MR. Anatomical variations in the level of origin of the sigmoid colon from the descending colon and the attachment of the sigmoid mesocolon. *Clin Anat* 2010; 23: 179–185 doi:10.1002/ca.20910
- [39] Moug SJ, Fountas S, Johnstone MS et al. Analysis of lesion localisation at colonoscopy: outcomes from a multi-centre U.K. study. *Surg Endosc* 2017; 31: 2959–2967
- [40] Cho YB, Lee WY, Yun HR et al. Tumor localization for laparoscopic colorectal surgery. *World J Surg* 2007; 31: 1491–1495 doi:10.1007/s00464-020-07443-5
- [41] Lam DT, Kwong KH, Lam CW et al. How useful is colonoscopy in locating colorectal lesions? *Surg Endosc* 1998; 12: 839–841 doi:10.1007/s004649900725
- [42] Tabibian N, Michaletz PA, Schwartz JT et al. Use of an endoscopically placed clip can avoid diagnostic errors in colonoscopy. *Gastrointest Endosc* 1988; 34: 262–264 doi:10.1016/s0016-5107(88)71326-7
- [43] Vignati P, Welch JP, Cohen JL. Endoscopic localization of colon cancers. *Surg Endosc* 1994; 8: 1085–1087 doi:10.1007/BF00705725
- [44] Piscatelli N, Hyman N, Osler T. Localizing colorectal cancer by colonoscopy. *Arch Surg* 2005; 140: 932–935 doi:10.1001/archsurg.140.10.932
- [45] Che K, Ye C, Yao Y et al. Deep learning-based biological anatomical landmark detection in colonoscopy videos. *arXiv preprint arXiv. 210802948*. 2021
- [46] Saito H, Tanimoto T, Ozawa T et al. Automatic anatomical classification of colonoscopic images using deep convolutional neural networks. *Gastroenterol Rep* 2021; 9: 226–233 doi:10.1093/gastro/goaa078
- [47] Yao H, Stidham RW, Gao Z et al. Motion-based camera localization system in colonoscopy videos. *Med Image Anal* 2021; 73: 102180 doi:10.1016/j.media.2021.102180
- [48] Ellul P, Fogden E, Simpson C et al. Colonic tumour localization using an endoscope positioning device. *Eur J Gastroenterol Hepatol* 2011; 23: 488–491 doi:10.1097/MEG.0b013e328346974b
- [49] Szura M, Pasternak A, Solecki R et al. Accuracy of preoperative tumor localization in large bowel using 3D magnetic endoscopic imaging: randomized clinical trial. *Surg Endosc* 2017; 31: 2089–2095 doi:10.1007/s00464-016-5203-4
- [50] Cheung HY, Chung CC, Kwok SY et al. Improvement in colonoscopy performance with adjunctive magnetic endoscope imaging: a randomized controlled trial. *Endoscopy* 2006; 38: 214–217
- [51] Houwen B, Hartendorp F, Giotis I et al. Computer-aided classification of colorectal segments during colonoscopy: a deep learning approach based on images of a magnetic endoscopic positioning device. *Scand J Gastroenterol* 2022; doi:10.1080/00365521.2022.2151320
- [52] Mahmood F, Durr NJ. Deep learning and conditional random fields-based depth estimation and topographical reconstruction from conventional endoscopy. *Med Image Anal* 2018; 48: 230–243 doi:10.1016/j.media.2018.06.005
- [53] Freedman D, Blau Y, Katzir L et al. Detecting deficient coverage in colonoscopies. *IEEE Trans Med Imaging* 2020; 39: 3451–3462 doi:10.1109/TMI.2020.2994221
- [54] Hwang SJ, Park SJ, Kim GM et al. Unsupervised monocular depth estimation for colonoscope system using feedback network. *Sensors* 2021; 21: 2691 doi:10.3390/s21082691
- [55] Armin MA, Barnes N, Alvarez J et al. Learning camera pose from optical colonoscopy frames through deep convolutional neural network (CNN). In: Cardoso MJ, Arbel T, Luo X (eds.) *Computer Assisted and Robotic Endoscopy and Clinical Image-Based Procedures*. Cham: Springer International Publishing; 2017: 50–59
- [56] Zhang S, Zhao L, Huang S et al. A template-based 3D reconstruction of colon structures and textures from stereo colonoscopic images. *IEEE Transactions Med Robotics Bionics* 2021; 3: 85–95
- [57] Walluscheck S, Wittenberg T, Bruns V et al. Partial 3D-reconstruction of the colon from monoscopic colonoscopy videos using shape-from-motion and deep learning. *Current Directions in Biomedical Engineering* 2021; 7: 335–338
- [58] Hong D, Tavanapong W, Wong J et al. 3D Reconstruction of virtual colon structures from colonoscopy images. *Comput Med Imaging Graph* 2014; 38: 22–33
- [59] Ma R, Wang R, Pizer S et al. Real-time 3D reconstruction of colonoscopic surfaces for determining missing regions. In: Shen D, Liu T, Peters TM (eds.) *Medical Image Computing and Computer Assisted Intervention – MICCAI 2019*. Cham: Springer International Publishing; 2019: 573–582
- [60] Hassan C, Pickhardt PJ, Kim DH et al. Systematic review: distribution of advanced neoplasia according to polyp size at screening colonoscopy. *Aliment Pharmacol Ther* 2010; 31: 210–217 doi:10.1111/j.1365-2036.2009.04160.x
- [61] Itoh H, Roth HR, Lu L et al. Towards automated colonoscopy diagnosis: binary polyp size estimation via unsupervised depth learning. In: Frangi AF, Schnabel JA, Davatzikos C (eds.) *Medical Image Computing and Computer Assisted Intervention – MICCAI 2018*. Cham: Springer International Publishing; 2018: 611–619
- [62] Itoh H, Oda M, Jiang K et al. Binary polyp-size classification based on deep-learned spatial information. *Int J Comput Assist Radiol Surg* 2021; 16: 1817–1828 doi:10.1007/s11548-021-02477-z
- [63] Su R, Liu J, Wu B et al. Accurate measurement of colorectal polyps using computer-aided analysis. *Eur J Gastroenterol Hepatol* 2021; 33: 701–708 doi:10.1097/MEG.0000000000002162
- [64] Abdelrahim M, Saiga H, Maeda N et al. Automated sizing of colorectal polyps using computer vision. *Gut* 2022; 71: 7–9 doi:10.1136/gutjnl-2021-324510
- [65] Kwak MS, Cha JM, Jeon JW et al. Artificial intelligence-based measurement outperforms current methods for colorectal polyp size measurement. *Dig Endosc* 2022; 34: 1188–1195 doi:10.1111/den.14318
- [66] Chadebecq F, Tilmant C, Bartoli A. How big is this neoplasia? live colonoscopic size measurement using the Infocus-Breakpoint Med Image Anal 2015; 19: 58–74 doi:10.1016/j.media.2014.09.002

- [67] Yoshioka M, Sakaguchi Y, Utsunomiya D et al. Virtual scale function of gastrointestinal endoscopy for accurate polyp size estimation in real-time: a preliminary study. *J Biomed Opt* 2021; 26: 096002 doi:10.1117/1.JBO.26.9.096002
- [68] Visentini-Scarzanella M, Kawasaki H, Furukawa R et al. A structured light laser probe for gastrointestinal polyp size measurement: a preliminary comparative study. *Endosc Int Open* 2018; 6: E602–E609 doi:10.1055/a-0577-2798
- [69] Shimoda R, Akutagawa T, Tomonaga M et al. Estimating colorectal polyp size with a virtual scale endoscope and visual estimation during colonoscopy: Prospective, preliminary comparison of accuracy. *Dig Endosc* 2022; 34: 1471–1477
- [70] Williams JG, Pullan RD, Hill J et al. Management of the malignant colorectal polyp: ACPGBI position statement. *Colorectal Dis* 2013; 15: 1–38 doi:10.1111/codi.12262
- [71] Matsuda T, Saito Y, Nakajima T et al. Macroscopic estimation of submucosal invasion in the colon. *Techniq Gastrointest Endosc* 2011; 13: 24–32
- [72] Ikehara H, Saito Y, Matsuda T et al. Diagnosis of depth of invasion for early colorectal cancer using magnifying colonoscopy. *J Gastroenterol Hepatol* 2010; 25: 905–912 doi:10.1111/j.1440-1746.2010.06275.x
- [73] Update on the Paris classification of superficial neoplastic lesions in the digestive tract. *Endoscopy* 2005; 37: 570–578 doi:10.1055/s-2005-861352
- [74] Kudo S, Lambert R, Allen JI et al. Nonpolypoid neoplastic lesions of the colorectal mucosa. *Gastrointest Endosc* 2008; 68: S3–S47 doi:10.1016/j.gie.2008.07.052
- [75] Fonolla R, van der Zander QEW, Schreuder RM et al. Automatic image and text-based description for colorectal polyps using BASIC classification. *Artif Intell Med* 2021; 121: 102178 doi:10.1016/j.artmed.2021.102178
- [76] Hassan C, Pickhardt PJ, Rex DK. A resect and discard strategy would improve cost-effectiveness of colorectal cancer screening. *Clin Gastroenterol Hepatol* 2010; 8: 865–869 doi:10.1016/j.cgh.2010.05.018
- [77] Kominami Y, Yoshida S, Tanaka S et al. Computer-aided diagnosis of colorectal polyp histology by using a real-time image recognition system and narrow-band imaging magnifying colonoscopy. *Gastrointest Endosc* 2016; 83: 643–649 doi:10.1016/j.gie.2015.08.004
- [78] Mori Y, Kudo SE, Misawa M et al. Real-time use of artificial intelligence in identification of diminutive polyps during colonoscopy: a prospective study. *Ann Intern Med* 2018; 169: 357–366
- [79] Barua I, Wieszczy P, Kudo S-E et al. Real-time artificial intelligence-based optical diagnosis of neoplastic polyps during colonoscopy. *NEJM Evidence* 2022; 1: doi:10.1056/EVIDoA2200003
- [80] Garcia-Rodriguez A, Tudela Y, Cordova H et al. In vivo computer-aided diagnosis of colorectal polyps using white light endoscopy. *Endosc Int Open* 2022; 10: E1201–E1207 doi:10.1055/a-1881-3178
- [81] Hassan C, Balsamo G, Lorenzetti R et al. Artificial intelligence allows leaving-in-situ colorectal polyps. *Clin Gastroenterol Hepatol* 2022; 20: 2505–2513.e4
- [82] Houwen B, Hazewinkel Y, Giotis I et al. Computer-aided diagnosis for optical diagnosis of diminutive colorectal polyps including sessile serrated lesions: a real-time comparison with screening endoscopists. *Endoscopy* 2023; doi:10.1055/a-2009-3990
- [83] Minegishi Y, Kudo SE, Miyata Y et al. Comprehensive diagnostic performance of real-time characterization of colorectal lesions using an artificial intelligence-assisted system: a prospective study. *Gastroenterology* 2022; 163: 323–325 e323 doi:10.1053/j.gastro.2022.03.053
- [84] Rondonotti E, Hassan C, Tamanini G et al. Artificial intelligence-assisted optical diagnosis for the resect-and-discard strategy in clinical practice: the Artificial intelligence BLI Characterization (ABC) study. *Endoscopy* 2022; doi:10.1055/a-1852-0330
- [85] Leggett B, Whitehall V. Role of the serrated pathway in colorectal cancer pathogenesis. *Gastroenterology* 2010; 138: 2088–2100 doi:10.1053/j.gastro.2009.12.066
- [86] Ijspeert JE, Bastiaansen BA, van Leerdam ME et al. Development and validation of the WASP classification system for optical diagnosis of adenomas, hyperplastic polyps and sessile serrated adenomas/polyps. *Gut* 2016; 65: 963–970 doi:10.1136/gutjnl-2014-308411
- [87] Kamitani Y, Nonaka K, Isomoto H. Current status and future perspectives of artificial intelligence in colonoscopy. *J Clin Med* 2022; 11: 2923 doi:10.3390/jcm11102923
- [88] Choi K, Choi SJ, Kim ES. Computer-aided diagnosis for colorectal cancer using deep learning with visual explanations. In: 2020 42nd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC). 2020: 1156–1159
- [89] Yang YJ, Cho BJ, Lee MJ et al. Automated classification of colorectal neoplasms in white-light colonoscopy images via deep learning. *J Clin Med* 2020; 9: 1593 doi:10.3390/jcm9051593
- [90] Zorron Cheng Tao Pu L, Maicas G, Tian Y et al. Computer-aided diagnosis for characterization of colorectal lesions: comprehensive software that includes differentiation of serrated lesions. *Gastrointest Endosc* 2020; 92: 891–899
- [91] Choi SJ, Kim ES, Choi K. Prediction of the histology of colorectal neoplasm in white light colonoscopic images using deep learning algorithms. *Sci Rep* 2021; 11: 5311 doi:10.1038/s41598-021-84299-2
- [92] Gong EJ, Bang CS, Lee JJ et al. No-code platform-based deep-learning models for prediction of colorectal polyp histology from white-light endoscopy images: development and performance verification. *J Pers Med* 2022; 12: 963 doi:10.3390/jpm12060963
- [93] Meng S, Zheng Y, Wang W et al. A computer-aided diagnosis system using white-light endoscopy for the prediction of conventional adenoma with high grade dysplasia. *Dig Liver Dis* 2022; 54: 1202–1208
- [94] Gupta N, Bansal A, Rao D et al. Prevalence of advanced histological features in diminutive and small colon polyps. *Gastrointest Endosc* 2012; 75: 1022–1030 doi:10.1016/j.gie.2012.01.020
- [95] Ponugoti PL, Cummings OW, Rex DK. Risk of cancer in small and diminutive colorectal polyps. *Dig Liver Dis* 2017; 49: 34–37 doi:10.1016/j.dld.2016.06.025
- [96] Vleugels JLA, Hassan C, Senore C et al. Diminutive polyps with advanced histologic features do not increase risk for metachronous advanced colon neoplasia. *Gastroenterology* 2019; 156: 623–634.e623
- [97] Vleugels JLA, Koens L, Dijkgraaf MGW et al. Suboptimal endoscopic cancer recognition in colorectal lesions in a national bowel screening programme. *Gut* 2020; 69: 977–980 doi:10.1136/gutjnl-2018-316882
- [98] Meulen LWT, van de Wetering AJP, Debeuf MPH et al. Optical diagnosis of T1 CRCs and treatment consequences in the Dutch CRC screening programme. *Gut* 2020; 69: 2049–2051 doi:10.1136/gutjnl-2019-320403
- [99] Hayashi N, Tanaka S, Hewett DG et al. Endoscopic prediction of deep submucosal invasive carcinoma: validation of the narrow-band imaging international colorectal endoscopic (NICE) classification. *Gastrointest Endosc* 2013; 78: 625–632 doi:10.1016/j.gie.2013.04.185
- [100] Hosotani K, Imai K, Hotta K et al. Diagnostic performance for T1 cancer in colorectal lesions ≥ 10 mm by optical characterization using magnifying narrow-band imaging combined with magnifying chromoendoscopy: implications for optimized stratification by Japan Narrow-band Imaging Expert Team classification. *Dig Endosc* 2021; 33: 425–432

- [101] Takeda K, Kudo SE, Mori Y et al. Accuracy of diagnosing invasive colorectal cancer using computer-aided endocytoscopy. *Endoscopy* 2017; 49: 798–802 doi:10.1055/s-0043-105486
- [102] Tamai N, Saito Y, Sakamoto T et al. Effectiveness of computer-aided diagnosis of colorectal lesions using novel software for magnifying narrow-band imaging: a pilot study. *Endosc Int Open* 2017; 5: E690–E694 doi:10.1055/s-0043-105490
- [103] Ito N, Kawahira H, Nakashima H et al. Endoscopic diagnostic support system for cT1b colorectal cancer using deep learning. *Oncology* 2019; 96: 44–50 doi:10.1159/000491636
- [104] Lui TKL, Wong KKY, Mak LLY et al. Endoscopic prediction of deeply submucosal invasive carcinoma with use of artificial intelligence. *Endosc Int Open* 2019; 7: E514–E520
- [105] Nakajima Y, Zhu X, Nemoto D et al. Diagnostic performance of artificial intelligence to identify deeply invasive colorectal cancer on non-magnified plain endoscopic images. *Endosc Int Open* 2020; 8: E1341–E1348 doi:10.1055/a-1220-6596
- [106] Lu Z, Xu Y, Yao L et al. Real-time automated diagnosis of colorectal cancer invasion depth using a deep learning model with multimodal data (with video). *Gastrointest Endosc* 2022; 95: 1186–1194.e3
- [107] Luo X, Wang J, Han Z et al. Artificial intelligence-enhanced white-light colonoscopy with attention guidance predicts colorectal cancer invasion depth. *Gastrointest Endosc* 2021; 94: 627–638 e621
- [108] Tokunaga M, Matsumura T, Nankinzan R et al. Computer-aided diagnosis system using only white-light endoscopy for the prediction of invasion depth in colorectal cancer. *Gastrointest Endosc* 2021; 93: 647–653 doi:10.1016/j.gie.2020.07.053
- [109] Moons LMG, Bastiaansen BAJ, Richir MC et al. Endoscopic intermuscular dissection for deep submucosal invasive cancer in the rectum: a new endoscopic approach. *Endoscopy* 2022; 54: 993–998 doi:10.1055/a-1748-8573
- [110] Brand M, Troya J, Krenzer A et al. Development and evaluation of a deep learning model to improve the usability of polyp detection systems during interventions. *United European Gastroenterol J* 2022; 10: 477–484
- [111] Kudo SE, Ichimasa K, Villard B et al. Artificial intelligence system to determine risk of T1 colorectal cancer metastasis to lymph node. *Gastroenterology* 2021; 160: 1075–1084 e1072 doi:10.1053/j.gastro.2020.09.027
- [112] Ichimasa K, Kudo SE, Mori Y et al. Artificial intelligence may help in predicting the need for additional surgery after endoscopic resection of T1 colorectal cancer. *Endoscopy* 2018; 50: 230–240 doi:10.1055/s-0043-122385
- [113] Takamaru H, Stammers M, Yanagisawa F et al. Conditional inference tree models to perceive depth of invasion in T1 colorectal cancer. *Surg Endosc* 2022; 36: 9234–9243 doi:10.1007/s00464-022-09414-4
- [114] Mori Y, Kudo SE, East JE et al. Cost savings in colonoscopy with artificial intelligence-aided polyp diagnosis: an add-on analysis of a clinical trial (with video). *Gastrointest Endosc* 2020; 92: 905–911 e901
- [115] Wadhwa V, Alagappan M, Gonzalez A et al. Physician sentiment toward artificial intelligence (AI) in colonoscopic practice: a survey of US gastroenterologists. *Endosc Int Open* 2020; 8: E1379–E1384 doi:10.1055/a-1223-1926