

Robotically Assisted Microsurgery: Development of Basic Skills Course

Philippe André Liverneaux¹, Sarah Hendriks¹, Jesse C Selber², Sijo J Parekattil³

¹Department of Hand Surgery, Strasbourg University Hospital, Illkirch, France; ²Plastic Surgery, Anderson Cancer Center, The University of Texas, Houston, TX; ³Robotic Surgery and Urology, Winter Haven Hospital and University of Florida, Florida, FL, USA

Robotically assisted microsurgery or telemicrosurgery is a new technique using robotic telemanipulators. This allows for the addition of optical magnification (which defines conventional microsurgery) to robotic instrument arms to allow the microsurgeon to perform complex microsurgical procedures. There are several possible applications for this platform in various microsurgical disciplines. Since 2009, basic skills training courses have been organized by the Robotic Assisted Microsurgical and Endoscopic Society. These basic courses are performed on training models in five levels of increasing complexity. This paper reviews the current state of the art in robotically assisted microsurgical training.

Keywords Robotics / Microsurgery / Computer simulation / Training

Correspondence:

Philippe André Liverneaux
Department of Hand Surgery,
Strasbourg University Hospital, 10 av
Baumann, F-67403 Illkirch, France
Tel: +33-688894779
Fax: +33-388552363
E-mail: philippe.liverneaux@
chru-strasbourg.fr

Supported by RAMSES: The
Robotic Assisted Microsurgical and
Endoscopic Society ([http://www.
roboticmicrosurgeons.org/publications/](http://www.roboticmicrosurgeons.org/publications/)).

No potential conflict of interest relevant
to this article was reported.

Received: 2 Feb 2013 • Revised: 15 Feb 2013 • Accepted: 18 Feb 2013
pISSN: 2234-6163 • eISSN: 2234-6171 • <http://dx.doi.org/10.5999/aps.2013.40.4.320> • Arch Plast Surg 2013;40:320-326

INTRODUCTION

Developed in the 1960s, microsurgery is a surgical technique using optical magnification that allows the microsurgeon to perform delicate movements which are difficult or impossible using the naked eye. Over the last few years, microsurgery has seen two major technical advances: supermicrosurgery [1] and telemicrosurgery [2]. The latter, telemicrosurgery or robotically assisted microsurgery can be defined as the technique of microsurgery that uses robotic telemanipulators to scale down surgical gestures or movements. The da Vinci robot (Intuitive Surgical Inc., Sunnyvale, CA, USA) is the only medical robot currently available on the market for robotically assisted microsurgery.

Several international groups, including the Robotic Assisted Microsurgical and Endoscopic Society (RAMSES, [http://www.](http://www.roboticmicrosurgeons.org)

[roboticmicrosurgeons.org](http://www.roboticmicrosurgeons.org)), have reported their experiences in various disciplines such as urology, plastic surgery, Ear, Nose, Throat (ENT) surgery, ophthalmology, neurosurgery, hand surgery, and peripheral nerve surgery [3]. RAMSES not only aims to promote microsurgery with robotic manipulators, but also aims to develop a new concept: endoscopic microsurgery. It combines the properties of microsurgery, endoscopic surgery, and telesurgery. This would not only allow magnification of the view of the operating field, but also enable microsurgeon to scale down the gestures or movements of the operator's hands, while taking a minimally invasive approach.

While telemicrosurgery is still in its infancy, RAMSES has organized telemicrosurgery classes since 2009. The purpose of this paper is to describe the models used for training, evaluation methods, and the organization and proceedings of basic and

advanced telemicrosurgery training.

THE DA VINCI ROBOT

The only surgical telemanipulator currently available on the market is the da Vinci robot (Intuitive Surgical Inc.). It consists of three components: a mobile instrument cart with four articulated arms, an imaging cart, and a console for the surgeon to control the robotic arms (Fig. 1). The mobile cart contains four articulated robotic arms, three of which carry surgical instruments and a fourth arm that manipulates the digital stereoscopic camera to visualize the surgical field. Each of these arms has multiple joints providing three-dimensional movement of the surgical instruments and optics. The surgical tools have articulation that provides intracorporeal range of movement to 360°, called “EndoWrist” technology. The tools available vary: dissecting forceps, scissors, scalpel, spreaders, etc. The fourth arm, which controls the optics, has a stereoscopic, high definition, endoscopic camera. The stereoscopic camera lens comprises a video imaging column similar to that used in conventional laparoscopy or arthroscopy and two light sources and dual stereoscopic cameras for three-dimensional vision with progressive magnification up to 12 to 15 times. The surgeons’ console is equipped with an optical viewing system, two telemanipulation handles, and five pedals. The optical viewing system, called the stereo viewer, offers a three-dimensional view of the operating field and displays text messages and icons that reflect the status of the system in real time. The two telemanipulation handles allow remote manipulation of the four articulated robotic arms. In its latest version, the da Vinci SI, the robot is equipped with 2 surgeon consoles to allow for the simultaneous use with two operators: the primary robotic surgeon and a surgical assistant. In this mode, the 3 robotic arms can be utilized at the same time by the two operators.

Five properties of the da Vinci robot are essential in telemicro-

surgery. The optical magnification of the operating field is obtained by the optical and digital magnification of the stereoscopic camera. The suppression of physiological tremor improves the quality of surgical movements. The scaling down of surgical movements improves accuracy by reducing the surgeon’s movements by a factor of 3 (position “fine”) or 5 (position “extra-fine”). The ergonomic design of the surgeons’ console is very useful in microsurgery because it improves the comfort of the surgical movements by simplifying the motion. The possibility of minimally invasive surgery allows the microsurgeon to work in unique operative fields with minimal cutaneous incisions.

Robotic telemanipulators have often been criticized for not having tactile feedback. In reality, it has been clearly demonstrated that force or tactile feedback is not absolutely necessary in microsurgery: first, because the range of motion in the microsurgical field is minimal, and second, because the perception of a 9/0 and 10/0 nylon yarn voltage is at the limit of human physiology [4]. Successful microsurgery with suture-assisted microsurgical robots has already been reported [5]. In addition, a new robotic platform, the Amadeus telemanipulator (Titan), that will be available soon, will be equipped with tactile feedback. It is not difficult to imagine that, in the future, there will be a robot able to replicate microsurgical maneuvers with tactile feedback force, possibly enabling endoscopic supermicrosurgery [1].

TRAINING MODELS

Among the models currently available in conventional microsurgery, there are nonliving non-biological models (latex, silicone, Gore-Tex, PracticeRat), nonliving biological models (artery of the chicken wing, pig’s feet, placenta) and living models (mouse, rat, rabbit) [6,7]. The ideal model would meet the following specifications: inexpensive, easily available, similar to the biological tissue, without ethical issues [8]. No model perfectly fulfills all these specifications and full education in telemicrosurgery must have several levels of increasing complexity, the ability to address technical challenges, and applicability to varying microsurgical procedures.

The first level is to become familiar with the robot and master the basic skills of telemicrosurgery. This can be done either by means of an analog simulator such as plastic rings (Fig. 2) or by means of a virtual simulator. Several simulators are available on the market: Mimic, Ross [9], dV-Trainer [10], and da Vinci skills simulator [11]. The virtual simulators assess students’ performances in terms of time to complete the exercise, gesture accuracy, missed targets, instrument collision, drops, etc. This is one reason why virtual simulators allow for highly efficient self-teaching (Fig. 3).


Fig. 1. Da Vinci Robot

The robot da Vinci SI contains three parts (from right to left): a mobile cart with four articulated arms, an imaging cart, and a console for the surgeon to control the robot arm (reprinted with permission from Intuitive Surgical, Sunnyvale, CA, USA).




Fig. 2. Robot set-up

Installation of a robot da Vinci SI for level 1 training. Plastic rings must be manipulated and moved from one display stand to the other. No measure of performance is possible with this model, except the time of realization of the task.

**Fig. 3. Simulator set-up**


Installation of the da Vinci skills simulator for level 1 training. Several tasks of increasing difficulty can be developed, including vascular sutures. Precise measures of performance allow for self-education.




The second level is to use telemicrosurgical sutures on inexpensive models that pose no ethical dilemma in terms of animal testing. We use calibrated earthworms (*Lombicus rubellus*) with an average length of 60 mm and a mean outer diameter of 3 mm (Fig. 4). Each worm is anaesthetized by soaking in a 1% xylocaine solution until a significant slowdown in locomotor activity is observed. Each end of the worm is then cut with a scalpel followed by gentle finger pressure applied from one end to the other in order to completely eviscerate the worm. One obtains a hollow tube, the lumen, which replicates the lumen of a vessel of an average internal diameter of 2 mm. A frank cut in the middle of the model produces two segments of equal length ready to be anastomosed [12]. After anastomosis, the permeability and tightness of the model can be tested by injecting saline solution through one end [13]. It is also possible to use the femoral vessels of chicken thighs (Fig. 5). However, this model

Fig. 4. Earthworm model

Preparation of one worm for level 2 training. A terminoterminal telemicrosurgical anastomosis is realized with a 10/0 nylon thread.

**Fig. 5. Chicken artery model**

Preparation of a femoral artery of a chicken leg for level 2 training. A terminoterminal telemicrosurgical anastomosis is realized with a 10/0 nylon thread.



is less attractive than the worm because of its higher cost and the greater difficulty of its preparation.


The third level is to perform telemicrosurgical sutures on living models. The rat model for instance, must be used in accordance with current legislation on animal experimentation. The telemicrosurgeon, as well as the microsurgeon, must learn to perfectly master the vascular sutures: venous and arterial (Fig. 6) [14] and the nerve suture techniques (Fig. 7) [15].

The fourth level is to perform more complex telemicrosurgical maneuvers. A thawed fresh human cadaver is an ideal model on which to learn to manipulate nerve regrowth scaffolds (Fig. 8), and to make pedicled flaps (Fig. 9) [16] and free flaps [17]. Reimplantations on living animals can also be performed [18].


The fifth and final level is reserved for endoscopic telemicrosurgery. It consists of telemicrosurgical maneuvers through minimally invasive incisions. Reconstructive surgery of the brachial plexus [19] or the removal of large flaps of the latissimus dorsi

Fig. 6. Rat artery model


Preparation of an artery of the tail of a rat for level 3 training. A terminoterminal telemicrosurgical anastomosis is realized with a 10/0 nylon thread.

**Fig. 7. Rat nerve model**

Preparation of a rat sciatic nerve for level 3 training. A terminoterminal telemicrosurgical anastomosis is realized with a 10/0 nylon thread.

**Fig. 8. Cadaver nerve model**

Preparation of a fresh human cadaver forearm for level 4 training. A graft of a nerve median is realized with a nerve guide with a 7/0 nylon thread.

**Fig. 9. Cadaver flap model**

Preparation of a fresh human cadaver hand for level 4 training. The pedicle of the first dorsal interosseous space is isolated before raising a kite flap.



muscle or of the rectus abdominis muscle [20] are particularly suited to endoscopic telemicrosurgery.

METHODS OF ASSESSMENT

Despite advances in surgical training, microsurgical training is still based on an apprenticeship model. Considering the complexity of microsurgery and the consequences of failure, the lack of a standardized, quantitative system to evaluate surgeons' skill, provide feedback, and measure training endpoints raises major quality control issues.

The Structured Assessment of Microsurgical Skills (SAMS) is a model designed to assess technical skills during microsurgery that was formulated from other assessment tools such as the Imperial College Surgical Assessment Device (ICSAD) and the Observed Structured Assessment of Technical Skills (OSATS) [21-24]. In a recent study, we validated the SAMS instrument

and applied it to a cohort of microsurgical trainees, demonstrating acceptable inter-rater reliability and a reproducible learning curve [25].

Robotic microsurgical skill combines many of the same principles and specific skill techniques as conventional microsurgery, but has the additional skill set that is unique to the use of the surgical robot. In order to capture this blend of skills for robotic microsurgery, we combined a validated robotic surgical rating system [26,27] with the SAMS, to create the Structured Assessment of Robotic Microsurgical Skill (SARMS) (Table 1). This evaluation system combines two separately validated skills assessment systems into a single evaluation system that encompasses both robotic skills and microsurgical skills. The SARMS is in the process of being validated by testing inter- and intra-rater reliability in a laboratory model. This is the first evaluation system of its kind and will serve as a guide to robotic microsurgical

Table 1. Structured Assessment of Robotic Microsurgery Skills (SARMS)

Skills	1	3	5
Micro skills			
Dexterity			
Bimanual dexterity	Lack of use of non-dominant hand	Occasionally awkward use of non-dominant hand	Fluid movements with both hands working together
Tissue handling	Frequently unnecessary force with tissue damage	Careful but occasional inadvertent tissue damage	Consistently appropriate with minimal tissue damage
Visuo-spatial ability			
Micro suture placement	Frequently lost suture and uneven placement	Occasionally uneven suture placement	Consistently, delicately and evenly spaced sutures
Knot technique	Unsecure knots	Occasional awkward knot tying and improper tension	Consistently, delicately and evenly placed sutures
Operative flow			
Motion			
	Many unnecessary or repetitive moves	Efficient but some unnecessary moves	Economy of movement and maximum efficiency
Speed			
	Excessive time at each step due to poor dexterity	Efficient time but some unnecessary or repetitive moves	Excellent speed and superior dexterity without awkward moves
Robotic skills			
Camera movement			
	Unable to maintain focus or suitable view	Occasionally out of focus and inappropriately view	Continually in focus and appropriate view
Depth perception			
	Frequent inability to judge object distance	Occasional empty grasp	Consistently able to judge spatial relations
Wrist articulation			
	Little or awkward wrist movement	Occasionally inappropriate wrist movement or angles	Continually using full range of endowrist motion
Cutting			
	Erratic suture length or uncoordinated cutting effort	Occasionally long/short suture and occasional dis-coordinated cutting motion	Smooth cutting motion with consistent suture length
Fourth arm transition			
	Inability to change from needle driver to scissors without prompting	Occasionally awkward changes from needle driver to scissors	Consistent and fluid transition from needle driver to scissors
Communication			
	Always delayed requests for instrument reposition	Occasional inability to recognize out of position instrument	Consistent and timely requests for appropriate instruments
Atraumatic needle handling			
	Consistent bending/breakage of needle	Occasional bending/breakage of needle	Consistently undamaged needle
Atraumatic tissue handling			
	Consistent inappropriate grasping/crushing or over spreading of tissue	Occasional inappropriate grasping/crushing or over spreading of tissue	Consistently gentle handling of tissue

Overall	1	2	3	4	5
Overall performance	Poor	Borderline	Satisfactory	Good	Excellent
Indicative skill	Novice	Advanced beginner	Competent	Proficient	Expert

credentialing, as well as training endpoints for a robotic microsurgical curriculum. This will also serve as a research tool to help determine the learning curve for robotic microsurgical skills.

In addition to the SARMS, a demographic survey (Table 2) will be distributed in order to determine if years of training, level of training, gender, handedness, or previous experience with robotic, microsurgical, or laparoscopic training, or video games has any effect on robotic microsurgical skill acquisition.

ORGANIZATION OF BASIC COURSE SKILL

Since 2009, RAMSES has organized foundation telemicrosurgery courses, twice a year, at the IRCAD facility, Strasbourg, Europe. A basic course program will open soon in Florida, USA. Each course welcomes a maximum of 6 participants from dif-

ferent specialties: urology, plastic surgery, hand surgery, orthopedics, ophthalmology, and neurosurgery. Up to the present, 42 surgeons from 11 countries have attended the basic courses (Japan, Singapore, Brazil, USA, Belgium, France, Netherlands, Spain, Switzerland, Turkey, and Kuwait). Two to three expert members of RAMSES supervise the participants. Each class lasts a full day. In the morning, seminars are given by engineers of the Intuitive Surgical Company and by RAMSES experts. The program includes the history of medical robotics, description of the robots, presentation of the telemicrosurgical models, and current clinical applications. The morning session ends at the laboratory, with a demonstration of the setup of the da Vinci robot by an engineer from the company Intuitive Surgical. The afternoon is devoted to a practical workshop. For a configuration of six participants, participants are divided into 3 groups of 2 people; 1 group of level 1 on the simulator and 2 groups of

Table 2. Robotic-Assisted Microvascular Anastomosis Survey

Name: <input style="width: 300px;" type="text"/>		No: <input style="width: 80px;" type="text"/>				
Please complete the following questionnaire by checking the appropriate box.						
Gender Female <input type="checkbox"/> Male <input type="checkbox"/>						
Age 25-30 <input type="checkbox"/> 31-35 <input type="checkbox"/> 36-40 <input type="checkbox"/> 41-45 <input type="checkbox"/> 46-50 <input type="checkbox"/> >50 <input type="checkbox"/>						
Hand dominance: Right-handed <input type="checkbox"/> Left-handed <input type="checkbox"/> Ambidextrous <input type="checkbox"/>						
Level of training: Resident <input type="checkbox"/> Clinical Fellow <input type="checkbox"/> Research Fellow <input type="checkbox"/> Junior Faculty <input type="checkbox"/> Senior Faculty <input type="checkbox"/>						
How much experience do you have with the following?		Novice	Advanced beginner	Competent	Proficient	Expert
1.	Microsurgery experience	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
2.	Robotic experience	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
3.	Laparoscopic experience	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
4.	Video game experience	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Number of robotic simulator hours completed? None <input type="checkbox"/> <5 hr <input type="checkbox"/> 6-10 hr <input type="checkbox"/> >10 hr <input type="checkbox"/>						

level 2 or 3 on the robot. The first group goes to the console of a da Vinci skills simulator to perform exercises at level 1. The 2 other groups work with each of the two da Vinci SI robots, each equipped with 2 consoles: 1 main operator console and 1 assistant console. Level 2 and 3 exercises can be done during the basic course. A certificate is issued to each participant who has completed the training.

Advanced courses on freshly thawed cadavers are planned on demand to acquire levels 4 and 5. The type of procedures performed depends on the participant’s specialty according to their desire to develop a particular surgical technique.

In conclusion, the development of a basic course in telemicrosurgery or robotically assisted microsurgery is essential to ensure the credibility of this new discipline, to train future users, and develop new interventions.

REFERENCES

1. Todokoro T, Koshima I. Supermicrosurgery. In: Liverneaux PA, Berner SH, Bednar MS, et al., editors. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013. p.189-94.
2. Liverneaux P, Nectoux E, Taleb C. The future of robotics in hand surgery. *Chir Main* 2009;28:278-85.
3. Liverneaux PA, Berner SH, Bednar MS, et al. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013.
4. Panchulidze I, Berner S, Mantovani G, et al. Is haptic feedback necessary to microsurgical suturing? Comparative study of 9/0 and 10/0 knot tying operated by 24 surgeons. *Hand Surg* 2011;16:1-3.
5. Taleb C, Nectoux E, Liverneaux PA. *Telemicrosurgery: a feasibility study in a rat model*. *Chir Main* 2008;27:104-8.

6. Seyhan T, Ozerdem OR. Microsurgery training on discarded abdominoplasty material. *Plast Reconstr Surg* 2006;117:2536-7.
7. Hino A. Training in microvascular surgery using a chicken wing artery. *Neurosurgery* 2003;52:1495-7.
8. Hosnuter M, Tosun Z, Savaci N. A nonanimal model for microsurgical training with adventitial stripping. *Plast Reconstr Surg* 2000;106:958-9.
9. Kesavadas T, Stegemann A, Sathyaseelan G, et al. Validation of Robotic Surgery Simulator (RoSS). *Stud Health Technol Inform* 2011;163:274-6.
10. Perrenot C, Perez M, Tran N, et al. The virtual reality simulator dV-Trainer[®] is a valid assessment tool for robotic surgical skills. *Surg Endosc* 2012;26:2587-93.
11. Kelly DC, Margules AC, Kundavaram CR, et al. Face, content, and construct validation of the Da Vinci Skills Simulator. *Urology* 2012;79:1068-72.
12. Mantovani Ruggiero G. Earthworms. In: Liverneux PA, Berner SH, Bednar MS, et al., editors. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013. p.53-7.
13. Ramdhian RM, Bednar M, Mantovani GR, et al. Microsurgery and telemicrosurgery training: a comparative study. *J Reconstr Microsurg* 2011;27:537-42.
14. Lee JY, Shin AY. Vessels. In: Liverneux PA, Berner SH, Bednar MS, et al., editors. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013. p.59-67.
15. Berner SH. Nerves. In: Liverneux PA, Berner SH, Bednar MS, et al., editors. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013. p.69-73.
16. Huat A, Facca S, Lebailly F, et al. Are pedicled flaps feasible in robotic surgery? Report of an anatomical study of the kite flap in conventional surgery versus robotic surgery. *Surg Innov* 2012;19:89-92.
17. Maire N, Naito K, Lequent T, et al. Robot-assisted free toe pulp transfer: feasibility study. *J Reconstr Microsurg* 2012; 28:481-4.
18. Taleb C, Nectoux E, Liverneux P. Limb replantation with two robots: a feasibility study in a pig model. *Microsurgery* 2009;29:232-5.
19. Mantovani Ruggiero G, Liverneux P. Brachial plexus repair. In: Liverneux PA, Berner SH, Bednar MS, et al., editors. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013. p.89-97.
20. Selber JC, Pedersen JC. Muscle flaps. In: Liverneux PA, Berner SH, Bednar MS, et al., editors. *Telemicrosurgery: robot assisted microsurgery*. Paris: Springer; 2013. p.145-57.
21. Balasundaram I, Aggarwal R, Darzi LA. Development of a training curriculum for microsurgery. *Br J Oral Maxillofac Surg* 2010;48:598-606.
22. Temple CL, Ross DC. A new, validated instrument to evaluate competency in microsurgery: the University of Western Ontario Microsurgical Skills Acquisition/Assessment instrument. *Plast Reconstr Surg* 2011;127:215-22.
23. Chan W, Niranjana N, Ramakrishnan V. Structured assessment of microsurgery skills in the clinical setting. *J Plast Reconstr Aesthet Surg* 2010;63:1329-34.
24. Chan WY, Matteucci P, Southern SJ. Validation of microsurgical models in microsurgery training and competence: a review. *Microsurgery* 2007;27:494-9.
25. Selber JC, Chang EI, Liu J, et al. Tracking the learning curve in microsurgical skill acquisition. *Plast Reconstr Surg* 2012; 130:551e-8e.
26. Dulan G, Rege RV, Hogg DC, et al. Developing a comprehensive, proficiency-based training program for robotic surgery. *Surgery* 2012;152:477-88.
27. Dulan G, Rege RV, Hogg DC, et al. Content and face validity of a comprehensive robotic skills training program for general surgery, urology, and gynecology. *Am J Surg* 2012; 203:535-9.